

NATIONAL INSTITUTE FOR HEALTH AND CARE EXCELLENCE

INTERVENTIONAL PROCEDURES PROGRAMME

Interventional procedure overview of sacrocolpopexy using mesh to repair vaginal vault prolapse

The vaginal vault is a structure formed at the top of the vaginal canal after surgery to remove the womb and the cervix. Vaginal vault prolapse happens when the upper part of the vagina slips down from its usual position. Sacrocolpopexy is an operation that aims to support for the pelvic organs in their natural position. This is achieved by attaching a piece of mesh, usually from the top and occasionally the front and back of the vagina, to a ligament in the pelvis at the base of the spine or to a bone at the bottom of the spine.

The National Institute for Health and Care Excellence (NICE) has prepared this interventional procedure (IP) overview to help members of the interventional procedures advisory committee (IPAC) make recommendations about the safety and efficacy of an interventional procedure. It is based on a rapid review of the medical literature and specialist opinion. It should not be regarded as a definitive assessment of the procedure.

Date prepared

This IP overview was prepared in June 2016 and updated in June 2017.

Procedure name

- Sacrocolpopexy using mesh for vaginal vault prolapse repair

Specialist societies

- British Association of Urological Surgeons (BAUS)
- Royal College of Obstetricians and Gynaecologists (RCOG)
- British Society of Urogynaecology (BSUG)

Description

Indications and current treatment

Vaginal vault prolapse is when the upper part of the vagina descends from its usual position, sometimes out through the vaginal opening. It is more common after hysterectomy.

Vaginal vault prolapse may occur on its own or together with a cystocele (when the bladder sags into the vagina), rectocele (when the front wall of the rectum bulges into the lower wall of the vagina) or enterocele (when the intestine bulges into the upper wall of the vagina).

It can affect quality of life by causing symptoms of pressure and discomfort, and by its effect on urinary, bowel and sexual function.

Treatment options for vaginal vault prolapse depend on the severity of the symptoms. Treatment is rarely indicated if there are no symptoms. Mild-to-moderate prolapse may be treated with conservative measures such as pelvic floor muscle training, electrical stimulation and biofeedback. Topical oestrogens and mechanical measures such as pessaries may also be used.

Surgery may be needed when the prolapse is severe. A number of different surgical procedures are available for repairing vaginal vault prolapse using vaginal or abdominal (open, laparoscopic or robotic) approaches. Some procedures involve using mesh to provide additional support.

What the procedure involves

Sacrocolpopexy using mesh to repair vaginal vault prolapse is done with the patient under general anaesthesia, using an open or laparoscopic abdominal approach. Mesh is attached to the longitudinal ligament of the sacrum, most often at the level of the sacral promontory. The mesh is then attached to the apex of the vagina and sometimes to the anterior or posterior vaginal wall.

The procedure can be combined with surgery for stress urinary incontinence, such as colposuspension or sub-urethral sling placement.

Several different types of meshes or grafts have been used for this procedure, including synthetic meshes (polypropylene), allografts (cadaveric fascia lata) and xenografts (porcine dermis or small intestinal mucosa).

Literature review

Rapid review of literature

The medical literature was searched to identify studies and reviews relevant to sacrocolpopexy using mesh for vaginal vault prolapse repair. The following databases were searched, covering the period from 1 July 2007 to 28 November 2016: MEDLINE, PREMEDLINE, EMBASE, Cochrane Library and other databases. Trial registries and the Internet were also searched. No language restriction was applied to the searches (see appendix C for details of search strategy). Relevant published studies identified during consultation or resolution that are published after this date may also be considered for inclusion.

The following selection criteria (table 1) were applied to the abstracts identified by the literature search. Where selection criteria could not be determined from the abstracts the full paper was retrieved.

Table 1 Inclusion criteria for identification of relevant studies

Characteristic	Criteria
Publication type	<p>Clinical studies were included. Emphasis was placed on identifying good quality studies.</p> <p>Abstracts were excluded where no clinical outcomes were reported, or where the paper was a review, editorial, or a laboratory or animal study.</p> <p>Conference abstracts were also excluded because of the difficulty of appraising study methodology, unless they reported specific adverse events that were not available in the published literature.</p> <p>Non-randomised studies with samples size smaller than 100 patients or a follow-up inferior to 24 months were also excluded.</p>
Patient	Female patients with vaginal prolapse.
Intervention/test	Open, laparoscopic or robotic sacrocolpopexy.
Outcome	Articles were retrieved if the abstract contained information relevant to the safety and/or efficacy.
Language	Non-English-language articles were excluded unless they were thought to add substantively to the English-language evidence base.

List of studies included in the IP overview

This IP overview is based on data from about 5,790 patients included in 4 systematic reviews and meta-analyses^{1-3,9}, 2 randomised control trials^{4,5}, 3 prospective case series⁶⁻⁸ and 1 case report¹¹.

Other studies that were considered to be relevant to the procedure but were not included in the main extraction table (table 2) have been listed in appendix A.

Table 2 Summary of key efficacy and safety findings on sacrocolpopexy using mesh to repair vaginal prolapse**Study 1 Maher C 2016****Details**

Study type	Systematic review and meta-analysis
Country	Australia
Recruitment period	Search date: inception to July 2015; Databases searched: Cochrane Incontinence Group's Specialised Register of controlled trials, which contains trials identified from the Cochrane Central Register of Controlled Trials (CENTRAL), MEDLINE, WHO ICTRP and hand searching of journals and conference proceedings (2012-15) and ClinicalTrials.gov (searched January 2016). Reference lists of relevant articles were also searched.
Study population and number	The review covered 3414 women in 30 RCTs, of these 16 trials (n= 1608) reported only on vaginal vault prolapse treated by sacrocolpopexy.
Age and sex	All trials reported age and parity. The mean age of women was between 60 and 70 years in all trials except in Anger 2014 and Rondini 2015 where the mean age was between 55 to 60 years. Median parity was less than 3 in all trials except Rondini 2015 with mean parity of 3.8.
Patient selection criteria	Inclusion criteria: RCTs on different types of surgeries (transvaginal or abdominal routes, repair with or without mesh or native tissue repair, extent of surgery included hysterectomy or uterine sparing, with or without incontinence surgery) for women with apical vaginal prolapse, with at least 6 months follow-up and at least 20 women in each arm were included. Types of apical prolapse included: uterine prolapse, vault prolapse (post-hysterectomy), unspecified vaginal prolapse (uterine and/or vault prolapse). Exclusion criteria: quasi-randomised studies, cross over studies were excluded.
Technique	Six trials (Benson 1996, Lim 2012, Lo 1998, Maher 2004, Maher 2011 and Rondini 2015) compared a vaginal-based apical prolapse repair with sacrocolpopexy for apical prolapse and randomised 583 women, of which 83% were post-hysterectomy. Two trials with 204 women compared different graft materials: Culligan 2005 compared polypropylene mesh (Trelex Boston,) with cadaveric fascia lata (Tutoplast, Mentor) and more recently Culligan 2013 polypropylene mesh (Pelvitex, Bard) with acellular collagen matrix porcine dermis (Pelvisoft, Bard). Four trials compared access routes for sacrocolpopexy: 2 trials with 120 women (Costantini 2013; Freeman 2013) compared SCP and LSC and Anger 2014 and Paraiso 2011 with 157 women compared LSC and RASC. Four trials evaluated the efficacy of performing continence surgery at the time of sacrocolpopexy including 544 women (Brubaker 2008, Costantini 2007, Costantini 2008 and Trabuco 2014). Three evaluated with and without colposuspension (Brubaker 2008, Costantini 2007 and Costantini 2008). Trabuco 2014 compared colposuspension with mid-urethral sling at the time of sacrocolpopexy.
Follow-up	Six trials reported median/mean follow-up of less than 1 year (Anger 2014, Costantini 2013, Culligan 2013, Lim 2012, Paraiso 2011 and Trabuco 2014). Two-year results were reported in 4 studies (Benson 1996, Lo 1998, Maher 2004 and Maher 2011). Three to 4-year outcomes were reported in 2 trials (Costantini 2008 and Rondini 2015), and four trials reported outcomes at greater than five years (Brubaker 2008, Costantini 2007 and Culligan 2005).
Conflict of interest/source of funding	The lead author is an author of two studies included in the review. No other authors have any conflicts of interest. Review was supported by the Cochrane Incontinence Review Group-supported by NIHR UK. Thirteen trials were at low risk of bias related to financial conflict of interest with risk being unclear in 16 trials and high in 1.

Analysis

Follow-up issues: Loss to follow-up ranged from 0 (Costantini 2008), to less than 10% (Anger 2014, Benson 1996, Culligan 2013, Maher 2004, Maher 2011). At 5 years Culligan 2005 reported a 46% loss to follow-up. Attrition rate not stated in Costantini 2013.

Study design issues: Cochrane review methods were used. GRADEPRO software was used to assess the overall quality of evidence. Where data was sufficiently similar, meta-analysis was done using a fixed effect model.

Twenty-three studies were rated as at low risk of attrition bias, 2 studies were rated as at high risk of attrition bias and 5 as at unclear risk.

In 5 trials, data were analysed on an intention-to-treat basis (Brubaker 2008, Culligan 2013, Maher 2004, Maher 2011, Paraiso 2011 and Rondini 2015).

Preoperative prolapse status was reported in all trials but equal distribution and severity of prolapse between groups was not specifically reported in Benson 1996.

Study population issues:

All trials included those with stage 2 or greater apical prolapse and abdominal intervention in all trials was an open sacrocolpopexy except for Maher 2011 where laparoscopic access to the abdomen was utilised and Lim 2012 where either a laparoscopic or open approach was performed.

Post-hysterectomy prolapse-only patients were included in Maher 2004; Maher 2011 and the remainder included both uterine and post-hysterectomy prolapse. Some outcomes included in this overview include patients having sacrocolpopexy with concomitant hysterectomy.

Other issues: When reviewing the literature included in this systematic review an unreported safety event (corneal abrasion) was identified in the paper by Paraiso 2011. This safety event was not reported by Maher 2016 but was included in this report for completion.

Overlap with other table 2 papers.

Key efficacy and safety findings

Efficacy	Safety
<p>n=3,414 (vaginal vault prolapse n=1,608)</p> <p>Blood transfusion rate</p> <p><u>There was no difference between SCP 3% (3/91) versus vaginal procedure 0/97</u></p> <p>RR 0.26, 95% CI 0.04 to 1.57; 3 RCTs, n = 277; $I^2 = 0\%$</p> <p><u>There was no difference between SCP 9% (3/32) versus SCP with colposuspension 9% (3/34)</u></p> <p>RR 0.94, 95% CI 0.20 to 4.33; 1 RCT (n=66);</p> <p>Subjective failure (Awareness of prolapse)</p> <p><u>Favours SCP 7% (10/139) versus vaginal procedure* 16% (22/138)¹</u></p> <p>RR 2.11, 95% CI, 1.06 to 4.21; 3 RCTs (n=277); $I^2=0\%$</p> <p><u>There may be no difference between SCP using mesh 3% (1/29) versus SCP using biological graft 10% (3/29)**³</u></p> <p>RR 0.33, 95%CI, 0.04 to 3.02; 1 RCT (n=58)</p> <p><u>There was no difference between SCP 31% (22/71) versus SCP with colposuspension 37% (27/73) (7 years follow-up)</u></p> <p>RR 1.19, 95% CI, 0.75 to 1.89; 1 RCT (n=144)</p> <p>Objective failure (Any recurrent prolapse)</p> <p><u>Favours SCP 19% (35/189) versus vaginal procedure 34% (68/201)¹</u></p> <p>RR 1.89, 95% CI, 1.33 to 2.70; 4 RCTs (n=390); $I^2=41\%$ (1-2 years follow-up)</p> <p><u>There was no difference between SCP using mesh 11% (10/87) versus SCP using biological graft 23% (20/86)²</u></p> <p>RR 0.49, 99% CI, 0.20 to 1.25; 2 RCTs (n=173); $I^2=48\%$</p> <p><u>There was no difference between SCP or RASC 10% (5/50) versus LSC 9% (4/46)³</u></p> <p>RR 0.87; 95%CI, 0.25 to 3.06; 2 RCTs (n=96)</p> <p><u>There was no difference between SCP 38% (14/37) versus SCP with colposuspension 30% (15/33)</u></p> <p>RR 1.20, 95% CI, 0.69 to 2.10; 1 RCT (n=70).</p> <p>Objective failure</p> <p>1. Anterior compartment</p> <p><u>Favours SCP 6% (6/102) versus vaginal procedure 24% (23/97)</u></p> <p>RR 4.02, 95% CI, 1.71 to 9.49; 2 RCTs (n = 199); $I^2 = 22\%$</p> <p>2. Apical compartment</p> <p><u>Favours SCP 2% (3/134) versus vaginal procedure 21% (29/141)</u></p> <p>RR 8.15, 95% CI, 2.71 to 24.49; 3 RCTs (n=275); $I^2 = 0\%$, $p<0.001$</p> <p>3. Posterior vaginal compartment</p> <p><u>Favours SCP 3% (3/99) versus vaginal procedure 12% (12/100)</u></p> <p>RR 3.43, 95% CI, 1.10 to 10.66; 2 RCTs (n=199); $I^2=0\%$</p> <p>POPQ scores</p> <p>1. Point Ba</p> <p><u>Favours SCP 49% (53/108) versus vaginal procedure 51% (55/108)</u></p> <p>MD 0.80 cm, 95% CI, 0.41 to 1.19; 1 RCT (n = 108)</p> <p><u>Favours SCP using mesh versus SCP using biological graft</u></p> <p>MD 0.80 cm, 95% CI, 0.20 to 1.40, 1 RCT (n=58)</p> <p><u>There was no difference between LSC 49% (38/78) versus RASC 51% (40/78)</u></p> <p>MD 0.05 cm; 95% CI, -0.31 to 0.41; 1 RCT (n=78)</p>	<p>Mesh erosion</p> <p><u>There was no difference between SCP 3% (8/283) versus vaginal procedures 4% (9/291)</u></p> <p>RR 1.13; 95% CI, 0.47 to 2.69, 6 RCTs (n=574); $I^2=28\%$</p> <p><u>There was no difference between SCP using mesh 3% (3/86) versus SCP using biological graft 1% (1/87)</u></p> <p>RR 2.35, 95% CI, 0.36 to 15.40; 2 RCTs (n=173); $I^2=0\%$ (1-5 years follow-up)</p> <p><u>There was no difference between SCP or RASC 2% (2/96) versus LSC 0/90</u></p> <p>RR 0.22, 95% CI, 0.01 to 4.40; 3 RCTs (n=186); $I^2 = 0\%$</p> <p>Bladder injury</p> <p><u>There was no difference between SCP 2% (4/244) versus vaginal procedures 1% (2/267)¹</u></p> <p>RR 0.57, 95% CI, 0.14 to 2.36; 5 RCTs (n = 511); $I^2=0\%$</p> <p><u>There was no difference between SCP using mesh 1/111 versus SCP using biological graft 0/113²</u></p> <p>RR 2.51, 95% CI, 0.10 to 60.13; 2 RCTs (n= 224); $I^2=0\%$</p> <p><u>There was no difference between SCP or RASC 2% (2/102) versus LSC 4% (4/97)</u></p> <p>RR 1.75, 95% CI, 0.43 to 7.14, 3 RCTs (n=199); $I^2=0\%$</p> <p>Bowel Injury</p> <p><u>There was no difference between SCP 1% (2/143) versus vaginal procedures 0.6% (1/163)</u></p> <p>RR 0.63, 95% CI, 0.12 to 3.23; 3 RCTs (n=306); $I^2=0\%$</p> <p><u>There was no difference between SCP or RASC 4% (2/56) versus LSC 0/52</u></p> <p>RR 0.36, 95% CI, 0.04 to 3.32; 2 RCTs (n=108); $I^2=0\%$</p> <p>Repeat surgery</p> <p>1. Repeat surgery for prolapse</p> <p><u>Favours SCP 6% (11/187) vaginal procedure 14% (28/196)¹</u></p> <p>RR 2.28, 95% CI, 1.20 to 4.32; 4 RCTs (n=383); $I^2=0\%$</p> <p><u>There was no difference between SCP using mesh 1% (1/87) versus SCP using biological graft 1% (1/86)²</u></p> <p>RR 1.00, 95% CI, 0.07 to 15.24; 2 RCTs (n=173); $I^2=0\%$, (1-5 years follow-up)</p> <p><u>There was no difference between SCP 8% (2/24) versus LSC 9% (2/23)²</u></p> <p>RR 1.04, 95% CI, 0.16 to 6.80; 1 RCT (n=47)</p> <p><u>There was no difference between SCP 5% (7/128) versus SCP with colposuspension 4% (5/128)</u></p> <p>RR 0.71, 95%CI, 0.24 to 2.15; 3 RCTs (n=256); $I^2=0\%$</p> <p>2. Repeat surgery for SUI</p> <p><u>There may be no difference between SCP 3% (6/189) vaginal procedures 6% (12/206)¹</u></p> <p>RR 1.87, 95% CI, 0.72 to 4.86; 4 RCTs (n=395); $I^2=0\%$</p> <p><u>There was no difference between SCP using mesh 3% (1/29) versus SCP using biological graft 0/29</u></p> <p>RR 3.00, 95% CI, 0.13 to 70.74; 1 RCT (n=58)</p> <p><u>There was no difference between SCP 5% (5/92) versus SCP with colposuspension 8% (7/91)</u></p> <p>RR 1.42, 95%CI, 0.47 to 4.30; 1 RCT (n=183)</p> <p>3. Repeat surgery for mesh exposure</p>

SCP 51% (165/322) versus SCP with colposuspension 49% (157/322) (favours SCP with colposuspension)

MD -0.40 cm, 95% CI, -0.62 to -0.18; 1 RCT, (n=322)

2. Point Bp

Favours SCP 49% (53/108) versus vaginal procedure 51% (55/108)

MD 0.77 cm, 95% CI, 0.38 to 1.16; 1 RCT (n=108)

There was no difference between SCP using mesh 50% (29/58) versus SCP using biological graft 50% (29/58)

MD -0.20 cm, 95% CI, -0.51 to 0.11, 1 RCT (n=58)

SCP or RASC 50% (63/125) versus LSC 50% (62/125) (favours LSC)

MD -0.40, 95% CI, -0.76 to -0.05; 2 RCTs (n=125); $I^2=0\%$

Favours SCP 51% (165/322) versus SCP with colposuspension 49% (157/322)

MD 0.30, 95% CI, 0.11 to 0.49; 1 RCT (n=322)

3. Point C

Favours SCP 49% (53/108) versus vaginal procedure 51% (55/108)

MD 0.50 cm, 95% CI, 0.11 to 0.88; 1 RCT (n=108)

There was no difference between SCP using mesh 50% (29/58) versus SCP using biological graft 50% (29/58)

MD 0.31 cm, 95% CI, -0.41 to 1.03; 1 RCT (n=58)

There was no difference between SCP or RASC 50% (99/197) versus LSC 50% (98/197)

MD 0.15 cm, 95% CI, -0.52 to 0.83; 3 RCTs (n=197); $I^2 = 0\%$

There was no difference between SCP 51% (165/322) versus SCP with colposuspension 49% (157/322)

MD 0.20 cm, 95% CI, -0.11 to 0.51; 1 RCT (n=322)

4. Total vaginal length

Favours SCP 49% (53/108) versus vaginal procedure 51% (55/108)

MD -0.89 cm, 95%CI, -1.29 to -0.50; 1 RCT (n=108)

There was no difference between SCP using mesh 50% (29/58) versus SCP using biological graft 50% (29/58)

MD -0.10 cm, 95% CI, -0.69 to 0.49, 1 RCT (n=58)

SUI

Favours SCP 17% (21/127) versus vaginal procedures 31% (42/136)¹

RR 1.86, 95% CI, 1.17 to 2.94; 3 RCTs (n=263); $I^2 = 0$

There may be no difference between SCP or RASC 5% (2/38) versus LSC 9% (3/35)

RR 1.63 95% CI, 0.29, 9.18, 1 RCT (n=73)

There was no difference between SCP 62% (93/151) versus SCP with colposuspension 49% (79/144) (4 to 7 years follow-up)

random-effects RR 1.13, 95%CI, 0.63 to 2.04; 3 RCTs (n=295); $I^2 = 70\%$

Quality of life

PISQ

There was no difference between SCP 49% (54/110) versus vaginal procedure 51% (56/110) (4 years follow-up)

MD -1.2, 95% CI, -4.35 to 1.95, 1 RCT (n=110)

P-QOL (0-100)

There may be no difference between SCP 49% (54/110) versus vaginal procedure 51% (56/110) (4 years follow-up)

MD 22.70, 95% CI, -7.53 to 52.93, 1 RCT (n=110)

There may be no difference between SCP 49% (23/47) versus LSC 51% (24/47)

MD 0.70, 95% CI, -19.14 to 20.54; 1 RCT (n=47)

PFIQ-7

There was no difference between SCP using mesh 50% (57/115) versus SCP using biological graft 50% (58/115)

There was no evidence of a difference between SCP versus vaginal procedures

RR 1.14; 95% CI 0.35 to 3.64; $I^2=48\%$; 5 RCTs (n=497)

There was no difference between SCP using mesh 2% (2/86) versus SCP using biological graft 1% (1/87)

RR 2.00, 95% CI, 0.19 to 20.86; 2 RCTs (n=173); $I^2=0\%$

Bladder function

1. De novo urge incontinence

There was no evidence of difference between SCP 21% (6/29) versus vaginal procedure 33% (11/33)

RR 1.61, 95% CI, 0.68 to 3.81; 1 RCT (n = 62)

2. De novo urinary voiding dysfunction

There was no evidence of difference between SCP 3% (1/38) versus vaginal procedure 3% (1/37)

RR 1.03, 95% CI, 0.07 to 15.82, 1 RCT (n=75)

Dyspareunia

Favours SCP 16% (7/45) versus vaginal procedures 36% (22/61)²

RR 2.53, 95% CI, 1.17 to 5.50; 3 RCTs (n=106), $I^2=43\%$

Corneal abrasion

LSC 0/33 versus RASC 3% (1/33), $p=0.99$

*vaginal procedure was either total vaginal mesh or vaginal sacrospinous colpopexy

**polypropylene mesh versus cadaveric fascia (Culligan 2005), or acellular matrix porcine dermis (Culligan 2013).

¹moderate quality evidence

²Low quality evidence

³very low-quality evidence

<p>MD -7.00, 95% CI, -29.48 to 15.48; 1 RCT (n=115)</p> <p><u>There was no difference between LSC 49% (38/78) versus RASC 51% (40/78)</u></p> <p>MD 21.00, 95% CI, -46.76 to 88.76; 1 RCT (n=78)</p> <p>PFDI-20</p> <p><u>Small advantage favours SCP 49% (54/110) versus vaginal procedure 51% (56/110) (4 years follow-up)</u></p> <p>MD 7.90 95% CI 0.70 to 15.10; 1 RCT (n=110)</p> <p><u>There was no difference between SCP using mesh 50% (57/115) versus SCP using biological graft 50% (58/115)</u></p> <p>MD -6.00, 95% CI, -25.75 to 13.75; 1 RCT(n=115)</p>	
<p>Abbreviations used: Ba, most distal position of the remaining upper anterior vaginal wall; Bp, most distal portion of the remaining upper posterior vaginal wall; Point C, most distal edge of cervix or vaginal cuff scar; CI, confidence interval; LSC, laparoscopic sacrocolpopexy; MD, mean difference; RASC, robot-assisted sacrocolpopexy; RCT, randomised controlled trial; PFDI-20, pelvic floor distress inventory; PFIQ-7, pelvic floor impact questionnaire; PISQ, pelvic organ prolapse/urinary incontinence sexual questionnaire; POPQ, pelvic organ prolapse quantification; P-QOL, prolapse quality of life questionnaire (0-good quality of life, 100-poor quality of life); RR, risk ratio; SCP, abdominal sacrocolpopexy; SUI, stress urinary incontinence.</p>	

Study 2 Siddiqui NY (2015)

Details

Study type	Systematic review and meta-analysis
Country	USA
Recruitment period	Studies published until 4 June 2012
Study population and number	n=1,176 (approximately) 13 Studies included in the evaluating anatomic success : <ul style="list-style-type: none"> • 5 RCTs • 1 prospective study • 7 retrospective non-randomised cohort studies Studies included in the analysis of adverse events: <ul style="list-style-type: none"> • 13 studies above • 5 short-term comparative studies • 61 non-comparative studies
Age and sex	Adult women. Mean ages not reported.
Patient selection criteria	Women with any degree of apical prolapse having surgical treatment were included; The intervention of interest was sacrocolpopexy (abdominal, laparoscopic or robotic) using mesh.
Technique	Comparative studies with a follow-up greater than 6 months were included in the primary analyses. Non-comparative studies and shorter follow-up ones were included in the analysis of adverse events.
Follow-up	RCTs - 1 to 2.5 years Non-randomised comparative studies - 6 months to 8.3 years
Conflict of interest/source of funding	The author is supported by award number K12-DK100024 from the national Institute of Diabetes and Digestive and Kidney Disease. This study was done by the Society of Gynaecologic Surgeons and Systematic Review groups. No conflicts of Interest declared.

Analysis

Follow-up issues: Shorter follow-up studies were included for adverse events analysis.

Study design issues:

- Meta-analyses were performed when there were at least 3 studies reporting the same outcome.
- Due to the large number of case series, only studies that included a minimum of 200 patients were used in the analysis of adverse events.
- Abstracts were screened by 2 independent reviewers. Disagreements were resolved by consensus or the judgement of a third reviewer.
- Studies methodological quality was assessed by using the Agency for Healthcare Research and Quality three category system (A – Good, B – Fair, C – Poor). Individual Outcomes were graded according to the GRADE system.
- The author mentions the use of PRISMA statement.
- Quality of evidence:
 - Anatomic outcomes after sacrocolpopexy – moderate quality
 - Difference in reoperation between sacrocolpopexy and native tissue vaginal repairs - very low quality.
 - Evidence about bowel and bladder symptoms - insufficient.
 - Post-operative sexual function – low quality.
- One RCT was only published as an abstract and wasn't included in the meta-analysis.
- 38 studies were excluded from the systematic review.

Study population issues: The intervention of interest is sacrocolpopexy but the analysis included some patients that had hysteropexy and cervicopexy procedures. Percentages reported by the author in outcomes not included in the synthesis. Author doesn't report overall subgroups' percentages.

For mesh sacrocolpopexy, the majority of comparative studies used an open abdominal approach. Overall results from robot-assisted sacrocolpopexy (RASC) or laparoscopic sacrocolpopexy (LSC) may differ from this analysis' reported outcomes. The number of participants recruited by the 61 non-comparative studies included in the analysis of adverse events was not reported by the author.

Other issues: Five studies (Maher 2004, Benson 1996, Roovers 2004, Lo 1998 and Rondini 2011) included in this systematic review and meta-analysis were also included in Maher 2013, Study 1 in Table 2.

Key efficacy and safety findings

Efficacy	Safety				
Objective failure <u>Stage 2 or greater in POP-Q</u> <ul style="list-style-type: none">Mesh SCP (132/177) versus native tissue vaginal repair (119/192): pooled OR 2.04; 95% CI, 1.12-3.72, I²=31%, Phet=0.23.^aLSC 17%(10/60) versus SSLF 0/51:p<0.01^b <p>Studies were too heterogeneous to pool and obtain results regarding sexual function, bowel and bladder function.</p>	<u>Reoperation</u> SCP 13%(6/46) versus SSLF 16%(7/43), p=0.67 ^c				
Adverse events in 18 comparative studies					
Adverse event	Studie s**	No. Studies (excluded)	OR (95%CI)	Events	
				Mesh	Native tissue
Dindo 1					
Ileus/SBO	All	7 (-1)	9.45 (3.39-26.4) *	2% (16/814)	<1% (2/780)
	RCTs	2	9.55 (1.31-69.4)	5% (4/86)	0/108
Nerve Injury ^{*a}	All	5	0.61 (0.18-2.05)	1% (4/514)	1% (7/743)
	RCTs	2	8.32 (1.15-60.3)	5% (4/75)	0/83
Dyspareunia ^{*b}	All	5	0.42 (0.25-0.72) *	5% (23/445)	12% (46/384)
	RCTs	3	0.14 (0.06-0.33)	1% (1/107)	25% (27/106)
Dindo 2					
Bleeding ^{*c}	All	12 (-1)	1.00 (0.63-1.59)	3% (43/1317)	2% (37/1863)
	RCTs	3	1.02 (0.20-5.14)	2% (3/123)	2% (3/128)
DVT/PE	All	4 (-2)	1.36 (0.14-13.7)	<1% (2/569)	<1% (1/599)
	RCTs	0	-	-	-
Infection ^{*d}	All	7 (-1)	2.01 (0.91-4.45)	3% (17/676)	1% (9/617)
	RCTs	4	1.98 (0.60-6.55)	4% (7/171)	2% (4/193)
Dindo 3a					
Mesh/suture complication	All	7	3.26 (1.62-6.56) *	4% (28/650)	1% (6/537)
	RCTs	3	7.72 (1.08-55.2)	3% (4/122)	0/131
Dindo 3b					
Reoperation	All	7	0.76 (0.28-1.09)	7% (46/615)	10% (51/511)
	RCTs	4	0.97 (0.33-2.88)	16% (25/153)	17% (29/168)
Urinary tract injury	All	8 (-2)	1.68 (0.79-3.55)	2% (20/1068)	1% (9/1108)
	RCTs	3	1.65 (0.28-9.65)	2% (3/134)	1% (2/154)
Bowel injury	All	10(-2)	0.91 (0.35-2.37)	1% (8/1219)	1% (10/1574)
	RCTs	3	0.57 (0.06-5.54)	1% (1/130)	1% (2/147)
Dindo 4a					
ICU admission ^{*e}	All	4 (-2)	4.64 (0.42-50.6)	1% (3/561)	0/506
	RCTs	0	-	-	-
Dindo 5					
Death	All	4 (-3)	0.14 (0.003-6.97)	0/503	<1% (1/582)
	RCTs	1	0.14 (0.003-6.97)	0/47	2% (1/47)
*p<0.001					
**Studies excluded from the meta-analysis, no events occurred in either group.					
^{*a} Or neuropathy					
^{*b} Or sexual dysfunction					
^{*c} Or haematoma or transfusion					
^{*d} Wound or pelvic/cuff infection					
^{*e} Cardiovascular or pulmonary event					
Adverse events in comparative and non-comparative studies					
AE	Mesh sacrocolpopexy		Native tissue vaginal repair		P

		% (95CI)	Range	No. studies	AE	% (95CI)	Range	No. studies	AE	
Dindo 1										
Ileus/SBO	2.7 (1.7-3.9)	0-12	24	3% (137/4168)	0.2 (0.1-0.6)	0-0.5	11	<1% (3/1449)	<0.01	
Nerve Injury* ^a	1.3 (0-3.7)	0-8.2	12	4% (96/2601)	4.5 (1.8-8.2)	0-46	16	5% (147/2813)	0.1	
Dyspareunia* ^b	7.3 (3-13)	0-39	15	12% (371/2986)	9.9 (5.2-16)	0-58	17	9% (200/2180)	0.48	
Dindo 2										
Bleeding* ^c	1.5 (1-2.1)	0-12	34	2% (128/6555)	2.9 (1.5-4.8)	0-20	34	5% (367/7044)	0.05	
DVT/PE	0.6 (0.2-1.2)	0-2.8	15	1% (46/4579)	0.1 (0-0.3)	0-0.83	15	<1% (8/4114)	0.03	
Infection* ^d	2.2 (1.2-3.4)	0-7.9	25	2% (114/5519)	1.8 (0-8.3)	0-55	19	12% (558/4743)	0.6	
Dindo 3a										
Mesh/suture complication	4.2 (3.2-5.4)	0-18	40	4% (348/7831)	0.4 (0-1.7)	0-7.8	11	1% (13/1169)	<0.001	
Dindo 3b										
Reoperation	5.4 (3.8-7.1)	0.32-25	31	5% (367/7218)	3.7 (2.0-5.9)	0-33	22	3% (114/3872)	0.28	
Urinary tract injury	1.5 (0.8-2.3)	0-9.8	34	2% (113/6894)	0.6 (0.2-1.1)	0-3.5	25	1% (46/5111)	0.05	
Bowel injury	0.3 (0.1-0.6)	0-4.7	31	1% (37/6642)	0.6 (0.2-1)	0-3.8	28	1% (47/5744)	0.33	
Dindo 4a										
ICU admission* ^e	2.1 (0-6.3)	0-24	14	6% (281/4233)	0.5 (0.1-1.2)	0-4.5	13	1% (27/3532)	0.11	
Dindo 5										
Death	0.2 (0.1-0.4)	0-2.4	13	<1% (6/3343)	0.1 (0-0.4)	0-2.1	14	<1% (12/4105)	0.61	
Abbreviations used: AE, adverse event; CI, confidence interval; DVT, deep vein thrombosis; GRADE, grades for recommendation, assessment, development and evaluation; ICU, intensive care unit; I ² , percentage of total variation across studies due to heterogeneity; LSC, laparoscopic sacrocolpopexy; PE, pulmonary embolism; POP-Q, pelvic organ prolapse quantification system; PRISMA, preferred reported items for systematic reviews; RCT, randomised control trial; RASC, robot-assisted sacrocolpopexy; SBO, small bowel obstruction; SCP, sacrocolpopexy; NR, not reported; Phet, P value for statistical heterogeneity; SSLF, sacrospinous ligament fixation.										

^aThree of the four RCTs included in the meta-analysis reported anatomic success for individual compartments. Meta-analysis pools all compartments. Study quality: B, C, A, B.

^bOne cohort study reported similar anterior and apical outcomes but more posterior wall recurrences after SCP compared with SSLF.

^cThere were highly inconsistent results from RCTs regarding reoperation rates.

^dThe highest quality study reported no significant difference in all-cause reoperation. Adverse events were classified in Dindo categories: Grade II – mesh extrusion treatable with local estrogen, Grade IIIa – excision with none or local anaesthesia; Grade IIIb – excision with general anaesthesia. The author defined all mesh extrusions or erosions as a Dindo grade III

Study 3 Serati M (2014)

Details

Study type	Systematic review and meta-analysis
Country	Italy
Recruitment period	Databased searched systematically for records published between 2000 to 2013, included papers published between 2006 and 2013
Study population and number	n= 1,488 patients from 27 studies <ul style="list-style-type: none"> - 17 single arm studies - 10 comparative studies (4 RASC versus SCP and 6 RASC versus LSC)
Age and sex	Adult women. Mean age not reported.
Patient selection criteria	All English language original reports describing more than 10 sacrocolpopexy procedures performed using robotic assistance.
Technique	RASC was compared with SCP or LSC. Outcomes and complications of RASC were reported.
Follow-up	Mean follow-up not reported
Conflict of interest/source of funding	None

Analysis

Follow-up issues:

- Studies reviewed by 2 independent researchers
- When data from original papers wasn't clearly interpretable, the corresponding authors were contacted by email.

Study design issues:

- Some studies included in this systematic review present results of women having sacrocolpopexy and a concomitant hysterectomy

Study population issues: About 38% of women treated by sacrocolpopexy had associated hysterectomy, 33% had anti-incontinence procedures. Sacrocolpopexy with hysterectomy is analysed in a separate piece of NICE guidance ([IPG 577](#)). Subgroup analyses of patients not having hysterectomy were reported when available.

Other issues: One study (Paraiso 2011) included in this systematic review and meta-analysis was also included in the paper by Maher, Study 1 Table 2.

Key efficacy and safety findings

Efficacy	Safety																																										
n=1,488 patients from 27 studies	Surgical related complications (n=1488)																																										
Success and patient satisfaction with RASC <table><tr><th>Outcome</th><th>Incidence</th><th>Success rate</th></tr><tr><td>Objective cure (apical prolapse)</td><td><1% (2/246)</td><td>97% - 100%</td></tr><tr><td>Objective cure (all compartments)^a</td><td>6% (66/1029)</td><td>84% - 100%</td></tr><tr><td>Reoperation</td><td>3% (23/687)</td><td></td></tr><tr><td>Subjective cure</td><td>- ^b</td><td></td></tr></table> <p>^aThis considered studies with follow-ups greater than 2 years.</p> <p>^bGreat heterogeneity in subjective cure reporting.</p> <p>Prolapse recurrence (all compartments)</p> <p>RASC versus LSC: OR: 0.75; 95%CI, 0.36-1.57 (n=444), I²=1%, p=0.40.</p>	Outcome	Incidence	Success rate	Objective cure (apical prolapse)	<1% (2/246)	97% - 100%	Objective cure (all compartments) ^a	6% (66/1029)	84% - 100%	Reoperation	3% (23/687)		Subjective cure	- ^b		<table><tr><th colspan="2">RASC associated^a (3% 48/1488)</th></tr><tr><td>Vaginotomy</td><td>1% (14/1488)</td></tr><tr><td>Bladder injury</td><td>2% (26/1488)</td></tr><tr><td>Ureteral injury</td><td><1% (1/1488)</td></tr><tr><td>Bowel injury</td><td><1% (4/1488)</td></tr><tr><th colspan="2">Postoperative complications^b (2% 20/1118)</th></tr><tr><td>Bowel obstruction</td><td><1% (5/1118)</td></tr><tr><td>Port site hernia</td><td><1% (6/1118)</td></tr><tr><td>Port site nerve entrapment</td><td><1% (1/1118)</td></tr><tr><td>Abscess</td><td><1% (3/1118)</td></tr><tr><td>Peritonitis due to bowel injury</td><td><1% (2/1118)</td></tr><tr><td>Vaginal cuff dehiscence</td><td><1% (1/1118)</td></tr><tr><td>Feeling of traction requiring repeated surgery</td><td><1% (2/1118)</td></tr></table> <p>^aSatava grade 2 and 3 complications. There was significant heterogeneity among studies (p<0.001). There was 1 case of suture and needle being lost requiring a 2cm incision for retrieval.</p> <p>^bSevere complications, Claven-Dindo grade ≥ 3a, no grade 4 or 5 complications reported.</p> <p>Intraoperative complications</p> <p>RASC versus LSC: OR: 1.05; 95%CI, 0.52-2.12 (n=443), I²=0%, p=0.94.</p> <p>Conversion</p> <p>RASC versus LSC: OR: 0.89; 95%CI, 0.25-3.19 (n=443), I²=0%, p=0.72.</p> <p>Postoperative complications (all grades)</p> <p>RASC versus LSC: OR: 1.85; 95%CI, 0.96-3.75 (n=350), I²=37%, p=0.18.</p> <p>Severe postoperative complications (grade≥3)</p> <p>RASC versus LSC: OR: 0.56; 95%CI, 0.36-2.83 (n=430), I²=24%, p=0.73.</p> <p>Mesh Erosion</p> <p>RASC versus LSC: OR: 1.82; 95%CI, 0.51-6.45 (n=438), I²=0%, p=0.86.</p> <p>The incidence of mesh erosion ranged from 0% and 8% but there was significant heterogeneity amongst studies (p<0.01).</p> <p>Possible risk factors include vaginotomy and concomitant total hysterectomy.</p> <p>One study comparing RASC and total hysterectomy with RASC with supracervical hysterectomy reported that total hysterectomy had an increased risk of mesh erosion (0% following supracervical versus 14% following total hysterectomy, p=0.008, LE 2b).</p> <p>Lightweight mesh could be considered a protective factor. Comparing patients with specific information available 3 mesh erosions (1%) of 275 patients that had lightweight polypropylene mesh versus 26 mesh erosions (3.6%) among 715 women who had standard weight polypropylene mesh (p=0.03, Odds Ratio:0.3, 95%CI, 0.08-0.97). This is reported as being highly susceptible to bias.</p>		RASC associated^a (3% 48/1488)		Vaginotomy	1% (14/1488)	Bladder injury	2% (26/1488)	Ureteral injury	<1% (1/1488)	Bowel injury	<1% (4/1488)	Postoperative complications^b (2% 20/1118)		Bowel obstruction	<1% (5/1118)	Port site hernia	<1% (6/1118)	Port site nerve entrapment	<1% (1/1118)	Abscess	<1% (3/1118)	Peritonitis due to bowel injury	<1% (2/1118)	Vaginal cuff dehiscence	<1% (1/1118)	Feeling of traction requiring repeated surgery	<1% (2/1118)
	Outcome	Incidence	Success rate																																								
	Objective cure (apical prolapse)	<1% (2/246)	97% - 100%																																								
	Objective cure (all compartments) ^a	6% (66/1029)	84% - 100%																																								
	Reoperation	3% (23/687)																																									
	Subjective cure	- ^b																																									
	RASC associated^a (3% 48/1488)																																										
	Vaginotomy	1% (14/1488)																																									
	Bladder injury	2% (26/1488)																																									
	Ureteral injury	<1% (1/1488)																																									
	Bowel injury	<1% (4/1488)																																									
	Postoperative complications^b (2% 20/1118)																																										
	Bowel obstruction	<1% (5/1118)																																									
	Port site hernia	<1% (6/1118)																																									
	Port site nerve entrapment	<1% (1/1118)																																									
Abscess	<1% (3/1118)																																										
Peritonitis due to bowel injury	<1% (2/1118)																																										
Vaginal cuff dehiscence	<1% (1/1118)																																										
Feeling of traction requiring repeated surgery	<1% (2/1118)																																										
Abbreviations used: CI, confidence interval; LE, level of evidence; LSC, laparoscopic sacrocolpopexy; OR, odds ratio; RASC, robot-assisted sacrocolpopexy; SCP, sacrocolpopexy.																																											

Study 4 Nygaard I (2013)

Details

Study type	RCT
Country	USA
Recruitment period	2002 to 2005
Study population and number	n= 215 (104 SCP and urethropexy versus 111 SCP only) Adult women seeking treatment for apical pelvic organ prolapse with uterine preservation.
Age and sex	Female, 61.9 (mean)
Patient selection criteria	Women completing the in-person 2 year CARE visit follow-up were recruited into the extended CARE study, 92% (215/233) were included.
Technique	The CARE RCT compared the outcomes of women that had sacrocolpopexy for POP with or without concomitant Burch urethropexy (prophylactic anti-incontinence procedure).
Follow-up	7 years (median)
Conflict of interest/source of funding	One of the authors reported serving as a consultant to Johnson & Johnson and Key Tech. One other author reported receiving research grants from Pelvalon, Astellas, University of California/Pfizer, Pfizer, and Xanodyne; and serving as consultant to Astellas (advisory board), GlaxoSmithKline, Uromedica, IDEO, Pfizer and Xanodyne. Another author reported serving as a consultant to Intuitive surgical.

Analysis

Follow-up issues: 34/215(16%) women were lost to follow-up at year 5, 55/215(26%) at year 7. Follow-up extended up to 9 years but year 8 and 9 were excluded from the analysis because of small numbers.

Study design issues:

Due to lack of funding some of the sites follow-up was stopped, decreasing follow-up rates and limiting the number of participants. The study wasn't powered to detect differences inferior to 15%.

The study surgeries were performed by 21 surgeons from 7 sites.

Study population issues: None

Other issues: none.

Key efficacy and safety findings

Efficacy				Safety									
n= 215 (104 SCP and urethropexy versus 111 SCP only)				Frequency of suture and mesh erosion in women enrolled in CARE and extended care <table><tr><th></th><th>2 years</th><th>7 years</th></tr><tr><td>Suture erosion</td><td>0.9% (3/322)</td><td>1.2% (4/322)</td></tr><tr><td>Mesh erosion</td><td>5.3% (17/322)</td><td>9.9% (32/322)</td></tr></table> <p>Erosion occurred with all types of mesh. Overall probability of mesh erosion at 6.18 years was 10.5% (95% CI, 6.8%16.1%) when right censoring time was the last clinic visit.</p> <p>When the right censoring time was either a clinic visit or a last telephone interview, probability of mesh erosion was 9.9% (95%CI, 6.5% to 15%).</p> <p>Of the 23 women with mesh erosion, 11 were in the urethropexy group and 12 in the control group.</p> <p>In the CARE and extended CARE sample 15 had excision in the operating room (13 vaginal and 2 abdominal), 4 were given oestrogen cream and 4 were asymptomatic.</p> <p>7 women in the urethropexy group and 13 in the no urethropexy group had either surgery or received a urethral bulking agent injection.</p> <p>7 women in the urethropexy group and 5 women in the control group had either surgery or pessary for POP</p> <p>By year 7 at least 16% (36/215) of women in the extended CARE had additional surgery related to pelvic floor disorders: 11 recurrent POP, 14 for SUI and 11 for mesh complications.</p>		2 years	7 years	Suture erosion	0.9% (3/322)	1.2% (4/322)	Mesh erosion	5.3% (17/322)	9.9% (32/322)
	2 years	7 years											
Suture erosion	0.9% (3/322)	1.2% (4/322)											
Mesh erosion	5.3% (17/322)	9.9% (32/322)											
Estimated probability of failure from parametric survival models 2 and 7 years after abdominal sacrocolpopexy.													
	2 years follow-up												
	Urethropexy	No urethropexy	Treatment difference (95% CI)										
Pelvic organ prolapse													
Symptomatic failure	0.14	0.12	0.026 (-0.032 to 0.087)										
Anatomic failure	0.09	0.09	-0.005 (-0.093 to 0.087)										
Composite* failure	0.22	0.18	0.035 (-0.101 to 0.164)										
Urinary incontinence													
Stress	0.44	0.61	-0.175 (-0.296 to -0.043)										
Overall	0.59	0.67	-0.082 (-0.203 to 0.041)										
	7 years follow-up												
	Urethropexy	No urethropexy	Treatment difference (95% CI)										
Pelvic organ prolapse													
Symptomatic failure	0.29	0.24	0.049 (-0.060 to 0.162)										
Anatomic failure	0.27	0.22	0.05 (-0.161 to 0.271)										
Composite* failure	0.48	0.34	0.134 (-0.096 to 0.322)										
Urinary incontinence													
Stress	0.62	0.77	-0.154 (-0.266 to -0.037)										
Overall	0.75	0.81	-0.064 (-0.161 to 0.032)										
*anatomic or symptomatic failure													

Abbreviations used: CARE, colpopexy reduction efforts; CI, confidence interval LSC, laparoscopic sacrocolpopexy; POP, pelvic organ prolapse; RCT, randomised control trial; RSC, robotic sacrocolpopexy; SCP, sacrocolpopexy.

Study 5 Tate SB (2010)

Details

Study type	RCT
Country	USA
Recruitment period	2001 to 2003
Study population and number	n=100 (29/54 polypropylene mesh versus 29/46 fascia lata group at follow-up)
Age and sex	Adult women, mean age 58 ± 9 years
Patient selection criteria	Women enrolled in the double-blinded RCT comparing polypropylene and cadaveric fascia lata for sacrocolpopexy and completing 1 year follow-up were suitable to be included in the 5 years follow-up.
Technique	After selecting sacrocolpopexy as a treatment for POP each patient was randomised to polypropylene mesh and cadaveric fascia lata and followed-up for 1 year. After completion of the 1 year follow-up women would have the opportunity of taking part in the 5 years follow-up.
Follow-up	5 years
Conflict of interest/source of funding	One of the authors is a consultant and a paid instructor of CR Bard. Another author is a consultant and a paid instructor of CR Bard, receives research support from Solace Therapeutics, receives research support from and is a consultant and paid instructor at Boston Scientific and is a consultant and a paid instructor for Intuitive Surgical.

Analysis

Follow-up issues:

- Only 58 (58%) of the 100 subjects returned for the 5 year visit, 29/54 from the polypropylene group and 20/46 from the fascia lata group. Eleven individuals (11%) returned only questionnaires and were excluded because didn't have POP-Q examination.
- Lost to follow-up was 31% but wasn't significantly different in either group: Fisher exact test, $p=0.42$). Rate of follow-up wasn't significantly different between prolapse-stage groups (Fisher's exact test, $p=0.53$). Individuals whose surgery was an anatomical success were more likely to follow up than year 1 anatomic failures; the differences in the follow-up rates were not significant (Fisher's exact test, $p=0.79$). These conclusions held for both groups. Year 1 successes were not significantly more likely to follow-up than year 1 failures for the mesh group ($p=0.32$) or the fascia group ($p=1.0$). Additionally the difference between successes follow-up at year 1 wasn't significantly different amongst groups (mesh $p=0.47$, fascia $p=0.28$). Tests were run on a small number of subjects but suggest demographic and clinical similarities on both followed-up and lost to follow-up cohorts, alleviating response bias.

Study design issues:

- A computerised blocked randomisation scheme was held with allocation being submitted in an opaque closed envelope. Only surgeons were not blinded to intervention.
- Data for both studies collected by a single, masked, clinical research nurse.
- After 1 year, the surgeon told the patients which graft material was used. This might have impacted on the description of symptoms.
- No validated instrument was used to collect the information about subjective symptoms of prolapse.

Study population issues: None.

Other issues: None

Key efficacy and safety findings

Efficacy							Safety
n=100 (29/54 polypropylene mesh versus 20/46 fascia lata group) Mean ± SD POP-Q measurements and mean POP-Q stage at year 1 to 5, within comparators and between comparators.							
	Fascia			Mesh			Fascia vs Mesh
	1 year (n=46)	5 years (n=29)	P value ^a	1 year (n=54)	5 years (n=29)	P value ^a	P value ^b
Aa	-1.9±1.2	-1.8±1.5	0.87	-2.5±0.8	-2.6±0.7	0.40	0.66
Ba	-1.9±1.2	-1.8±1.5	0.95	-2.5±0.8	2.6±0.7	0.19	0.46
C	-8.1±2.7	-7.8±1.4	0.01*	-9±1.2	-8.1±1.4	0.0006*	0.22
Gh	2.4±0.7	2.5±0.7	0.51	2.3±0.6	2.4±0.8	0.95	0.42
Pb	3.4±0.7	3.2±0.8	0.02	3.6±0.8	3.1±1	0.31	0.36
TVL	9.3±1	8.4±1.2	0.0007*	9.4±1	8.5±1.1	0.0006*	0.46
Ap	-2.7±0.6	-2.7±0.8	1.00	-2.9±0.3	-2.9±0.3	1.0	0.79
Bp	-2.7±0.6	2.7±0.8	1.00	-2.9±0.3	-2.9±0.3	1.0	0.79
Stage	1±0.9	1±1	0.88	0.6±0.7	0.5±0.6	0.61	0.66
^a Signed rank test for within treatment group comparisons							
^b Rank sum test for the treatment group comparisons							
*despite statistically significant these differences weren't clinically significant							
Success rates at 1 and 5 years follow-up							
Definition	1 year polypropylene mesh (6)	1 year Cadaveric fascia (6)	P value	5 year polypropylene mesh	5 year Cadaveric fascia	P	
Objective anatomic	41/45 (91%)	30/44 (68%)	0.007 ^a	27/29 (93%)	18/29 (62%)	0.02 ^b	
Clinical	NA	NA	-	28/29 (97%)	26/29 (90%)	0.61 ^b	
^a P value is from Chi-squared test							
^b P values are from Fisher's exact test							
Objective anatomic and clinical failure frequencies							
	Fascia lata (n=29)			Mesh (n=29)			
Subjective complaints	-			-			
Vaginal bulge	13% (4/29)			-			
Symptoms of prolapse	10% (3/29)			-			
Any POP-Q point>0	10% (3/29)			-			
POP-Q point C1/2 TVL	-			--			
Surgical re-treatment	3%(1/29)			3% (1/29)			
Failure by clinical definition ^a	10% (3/29)			-			
Failure by objective anatomic definition ^b	38% (11/29)			7% (2/29)			
^a Complaints of vaginal bulge or symptoms of prolapse and a POP-Q point>0 or POP-Q pointC>1/2 TVL							
^b POP-Q point≥-1 (≥stage 2)							
Abbreviations used: Aa, anterior vaginal wall; Ba, most distal position of the remaining upper anterior vaginal wall; C, most distal edge of cervix or vaginal cuff scar; Gh, genital hiatus; Pb, peritoneal body; TVL, total vaginal length; Ap, posterior vaginal wall 3 cm proximal to the hymen, Bp, most distal portion of the remaining upper posterior vaginal wall							
BPCI, confidence interval; NA, not applicable; POP, pelvic organ prolapse; POP-Q, Pelvic organ prolapse quantification system; RCT, Randomised control study.							
							Year 1 follow-up There were 2 graft erosions, 1 in each treatment group. The polypropylene mesh erosion occurred at the posterior wall and eroded to the rectum requiring bowel resection and formation of colostomy. The subject was lost to follow-up. The subject with fascia lata graft erosion was lost to follow-up between years 1 and 5. At the year 5 visit the fascial erosion persisted and the patient presented with post-coital spotting, dyspareunia, vaginal discharge and odour.
							Year 5 follow-up There was 1 additional erosion in the polypropylene mesh group. The subject had a 2x3 cm apical mesh erosion. Laparoscopic vaginal removal of the mesh was needed. Necrotising fasciitis was developed post-operatively at the umbilical port site. The patient had a long stay in hospital but fully recovered. There were 2 retreated subjects with documented cystocele repairs between original surgery and 5 year follow-up, one in each group.

Study 6 Sarlos D (2013)

Details

Study type	Prospective case series
Country	Switzerland
Recruitment period	Initial prospective case series recruited between 2003 and 2007 Long-term follow-up exam occurred between July and September 2011
Study population and number	n= 101 , adult women
Age and sex	Age not reported.
Patient selection criteria	101 cases of LSC for uterine and post-hysterectomy prolapse enrolled in a prospective cohort with 12-month follow-up. Five years after surgery 99 of the 101 women were invited to a follow-up.
Technique	Women that decided to complete the 5 years follow-up had clinical examination and were asked to fill out two questionnaires: the German version of the Kings Health Questionnaire and a validated German version of the POP questionnaire. The degree of prolapse was documented using the POP-Q classification
Follow-up	60 months (mean)
Conflict of interest/source of funding	None.

Analysis

Follow-up issues:

- Only 85 of the 101 (84%) women included in the initial cohort participated in the long-term follow-up. From these, 17 patients could not come to the outpatient clinic because of age and comorbidities or relocation. Full follow-up was obtained for 68 women. There were 14 patients lost to follow-up.

Study design issues:

- Data from the 68 patients completing full follow-up was used to calculate the objective cure rates. For the subjective cure rates data from all 85 women was used.
- Surgical procedures were performed by 2 senior urogynaecologists experienced in LSC.

Study population issues: None.

Other issues: None

Key efficacy and safety findings

Efficacy					Safety				
Objective and subjective cure rates after LSC					Safety events after LSC				
	12 months (n=99)		60 months (n=68)			12 months (n=99)		60 months (n=85)	
	Total number	%	Total number	%		Total number	%	Total number	%
Subjective cure rate ^a	97/99	98.0	81/85 (81/99)	95 (82)	De novo SUI	24	24	32	38
Objective cure rate	91/99	92	57/68 (57/99)	84 (58)	Surgery for post-operative SUI	15	15	16	19
Recurrence of anterior wall	6/99	6	6/68(37)	9 (37)	Post-operative constipation	1	1	4	5
Recurrence of posterior wall	2/99	2	4/68 (35)	6 (35)	Post-operative voiding disorders	8	8	11	13
Apical recurrence	0	0	1/68 (32)	1 (32)	De novo urge incontinence	2	2	7	8
QoL score ^d	9.1	-	8.3	-	Severe de novo dyspareunia	1/47	2 ^c	10/41	24
Objective cure after LSC ^b					Mesh erosion	1	1	2	3
POP-Q	Pre-operatively (n=99)	12 months (n=99)	60 months (n=68)		There were 2 post-operative mesh protrusions into the bladder, 1 case happened 12 months after surgery and another case 60 months. Both cases had incidental bladder incision during LSC. Protruded mesh was removed by laparoscopy with partial excision of the anterior mesh and reconstruction of the bladder.				
Aa	-1 (±1.8)	-2 (±0.4)	- 2 (±1.0)						
Ba	1(±2.3)	-2 (±1.5)	-2 (±1.5)						
C	-1 (±3.4)	-7 (±2)	-6 (±1.2)						
Ap	-2 (±1.3)	-3 (±0.6)	-3 (±0.6)						
Bp	-2 (±3.1)	-3 (±1.1)	-3 (±3.2)						
Abbreviations used: Aa, anterior vaginal wall; Ap, posterior vaginal wall 3 cm proximal to the hymen; Ba, most distal position of the remaining upper anterior vaginal wall; Bp, most distal portion of the remaining upper posterior vaginal wall; C, most distal edge of cervix or vaginal cuff scar; LSC, laparoscopic sacrocolpopexy; POP, pelvic organ prolapse; QoL, quality of life; POP-Q, pelvic organ prolapse quantification system; SUI, Stress urinary incontinence.									

^aIn parentheses: number and percentage if every dropout is counted as a failure

^bResults are given as mean and standard deviation in parenthesis

^cAs only 41 (60 months follow-up) and 47 (at 12 months follow-up) declared themselves as sexually active, 41 or 47 were taken as 100%)

^dQuality of life assessed using a visual analogue scale from 1 to 10. Pre-operatively the quality of life index was 5.6

Study 7 Linder BJ (2015)

Details

Study type	Prospective case series
Country	USA
Recruitment period	2002 and 2012
Study population and number	n=70 , adult women
Age and sex	67 years (Median) , IQR [59-74]
Patient selection criteria	84 consecutive patients having RASC at the Mayo Clinic (Rochester, USA). Patients were excluded if RASC was converted to SCP.
Technique	ASC was selected after counselling regarding best treatment options. Patients were given 10-point Likert scale as well as PFDI and PFIQ-7 questionnaires to answer.
Follow-up	72 months (Median) [IQR 39-144]
Conflict of interest/source of funding	None.

Analysis

Follow-up issues:

Study design issues:

- All women were treated by a fellowship-trained female pelvic reconstructive surgeon and a fellowship-trained minimally-invasive urologist.
- Some of the patients had a concomitant anti-incontinence procedure at the time of RASC (robot-assisted sacrocolpopexy).
- Some patients were excluded as they had RASC converted to SCP (n=14). This was because of inability to dissect secondary to scarring, dense abdominal adhesions and failure to progress during pre-sacral dissection.

Study population issues: Concomitant mid-urethral sling was carried out in 55 (79%) patients

Other issues: None

Key efficacy and safety findings

Efficacy	Safety																																												
n=70, adult women <table><tr><td colspan="2">Objective failure rate (n=70)</td></tr><tr><td>Repeated surgery^a</td><td>1 years = 2% 3 years = 5% 6 years= 10% (frequencies not reported)</td></tr><tr><td colspan="2">Subjective failure rate (n=40)^b</td></tr><tr><td>Would recommend procedure to relative or friend</td><td>55% (22/40)</td></tr><tr><td>Would probably recommend procedure to relative or friend</td><td>25% (10/40)</td></tr><tr><td>Overall satisfaction^c</td><td>10 [IQR 8-10]</td></tr><tr><td>Symptomatic improvement^d</td><td>9 [IQR 8-10]</td></tr></table> <p>^arepeated surgeries included 2 patients having anterior colporrhaphy and 1 patient treated by posterior colporrhaphy. There was 1 case of apical prolapse at 128 months (10 years) follow-up.</p> <p>^bMedian follow-up was 90 months [IQR 56-120 months].</p> <p>^cMedian response to the question “How successful has your treatment for prolapse been?”, on a Likert scale (0=not at all success, 10=very successful).</p> <p>^dMedian post-operative symptomatic improvement, evaluated on a Likert scale, (0=much worse, 10=much better).</p> Further symptomatic follow-up* <table><tr><th>Scale</th><th>Score</th><th>Median [IQR]</th></tr><tr><td colspan="3">PFDI-20</td></tr><tr><td>POP Distress inventory-6</td><td>25-100</td><td>29.2 (25-37.5)</td></tr><tr><td>Colorectal-anal Distress Inventory-8</td><td>25-100</td><td>40.6 (28.1-47.7)</td></tr><tr><td>Urinary Distress Inventory-6</td><td>25-100</td><td>45 (35-60)</td></tr><tr><td colspan="3">PFIQ-7</td></tr><tr><td>Total Score</td><td>0-300</td><td>0 (0-28.6)</td></tr><tr><td>Bladder</td><td>0-100</td><td>0 (0-19)</td></tr><tr><td>Bowel</td><td>0-100</td><td>0 (0-4.8)</td></tr><tr><td>Pelvis</td><td>0-100</td><td>0 (0-4.8)</td></tr></table> <p>*Median follow-up for patients that did not complete a questionnaire at last follow-up was 49 months [IQR 8-93]</p>	Objective failure rate (n=70)		Repeated surgery ^a	1 years = 2% 3 years = 5% 6 years= 10% (frequencies not reported)	Subjective failure rate (n=40) ^b		Would recommend procedure to relative or friend	55% (22/40)	Would probably recommend procedure to relative or friend	25% (10/40)	Overall satisfaction ^c	10 [IQR 8-10]	Symptomatic improvement ^d	9 [IQR 8-10]	Scale	Score	Median [IQR]	PFDI-20			POP Distress inventory-6	25-100	29.2 (25-37.5)	Colorectal-anal Distress Inventory-8	25-100	40.6 (28.1-47.7)	Urinary Distress Inventory-6	25-100	45 (35-60)	PFIQ-7			Total Score	0-300	0 (0-28.6)	Bladder	0-100	0 (0-19)	Bowel	0-100	0 (0-4.8)	Pelvis	0-100	0 (0-4.8)	8.6% (6/70) had a total of 6 surgeries for recurrent prolapse or mesh complication. 2.7% (2/70) vaginal extrusion
Objective failure rate (n=70)																																													
Repeated surgery ^a	1 years = 2% 3 years = 5% 6 years= 10% (frequencies not reported)																																												
Subjective failure rate (n=40) ^b																																													
Would recommend procedure to relative or friend	55% (22/40)																																												
Would probably recommend procedure to relative or friend	25% (10/40)																																												
Overall satisfaction ^c	10 [IQR 8-10]																																												
Symptomatic improvement ^d	9 [IQR 8-10]																																												
Scale	Score	Median [IQR]																																											
PFDI-20																																													
POP Distress inventory-6	25-100	29.2 (25-37.5)																																											
Colorectal-anal Distress Inventory-8	25-100	40.6 (28.1-47.7)																																											
Urinary Distress Inventory-6	25-100	45 (35-60)																																											
PFIQ-7																																													
Total Score	0-300	0 (0-28.6)																																											
Bladder	0-100	0 (0-19)																																											
Bowel	0-100	0 (0-4.8)																																											
Pelvis	0-100	0 (0-4.8)																																											
Abbreviations used: IQR, Interquartile range; LSC, laparoscopic sacrocolpopexy; PFDI-20, pelvic floor distress inventory questionnaire; PFIQ-7, ; POP, pelvic organ prolapse; RASC, robot-assisted sacrocolpopexy; SCP, sacrocolpopexy;																																													

Study 8 Granese R (2009)

Details

Study type	Prospective case series
Country	Italy
Recruitment period	1999 to 2007
Study population and number	n= 165
Age and sex	Mean 67 years (range 58-76 years, SD 19.22)
Patient selection criteria	Women with diagnosis of vaginal vault prolapse between 2 nd and 4 th degree, according to HWS classification, that had LSC.
Technique	All women had a urogynaecological work-up before surgery, including a physical vulvovaginal examination* and an instrumental evaluation** with urodynamic investigation. During LSC mesh was always inserted laparoscopically.
Follow-up	Median 43 months (range 6-96 months)
Conflict of interest/source of funding	Not reported.

*Evaluation of descensus, urethra mobility test: Q-tip, perineal muscular balance, stress test/urinary incontinence test, proctologic evaluation, rectovaginal exploration in orthostatism, post-urination residual evaluation.

**Cystometry, uroflowmetry, valsalva leak point pressures, urethra pressure profile.

Analysis

Follow-up issues:

- At 1, 6, 12 months after surgery, a physical examination was carried out for all patients. After this period, women were contacted annually. Patients not able to be present in some of the follow-ups were contacted by phone and asked about the presence or absence of prolapse, urinary and bowel symptoms.
- There were 27 patients that were lost to follow-up: 4 died since surgery, 18 could not be contacted anymore and 5 declined to participate at follow-up.

Study design issues:

- None.

Study population issues:

Among the 165 women, 33 had already had other surgical procedures: 15 posterior colporrhaphy without relapse and 18 anterior colporrhaphy with 2 presenting a 2nd degree cystocele at the moment of LSC, and were therefore treated again. All additional corrections, except the enterocele and rectocele repairs were carried out after LSC. When necessary, a perineorrhaphy was also performed at the end of surgery.

Other issues: None

Key efficacy and safety findings

Efficacy			Safety
n=165			Symptoms after 8 years follow-up De novo urinary incontinence was present in 5% (7/138) of patients.
Objective cure rates (Median 43 months) Successful treatment was achieved in 95% (131/138) patients.			
Summary of symptoms before LSC and at follow-up			
	Before treatment	43 months (median)^a	
Vaginal vault prolapse symptoms			
Vaginal lump	83% (115/138) ^b	7% (10/138)	
Lower abdominal pain	69% (95/138) ^c	4% (6/138)	
Urinary symptoms			
Nycturia	17% (24/138)	4% (4/138)	
Dysuria	9% (13/138)	3% (4/138)	
Stress incontinence	4% (5/138)	7% (11/138)	
Mixed incontinence	17% (23/138)	14% (20/138)	
Pollakiuria	13% (18/138)	7% (10/138)	
Voiding dysfunctions	16% (22/138)	7% (9/138)	
Urge incontinence	11% (15/138)	18% (25/138)	
Recurrent urinary tract infections	16% (22/138)	5% (7/138)	
Bowel symptoms			
Constipation	7% (10/138)	13% (18/138)	
Obstructed defecation	1% (2/138)	6% (8/138)	
Urgency	0	2% (3/128)	
Pelvic pressure	67% (92/138)	9% (12/138)	
False urge to defecate	51% (70/138)	5% (7/138)	
Patient satisfaction			<p>*Recurrent vaginal prolapse was registered in 7 patients, in 3 the vaginal vault detached after a period of 7-20 days and was caused by the use of Vyprol (hence use of same has been suspended). In another 3 women (2 affected by 3rd degree prolapse and 1 by 2nd degree prolapse) the mesh detached after less than 1 month. These cases were performed by a less experienced surgeon and prolapse was associated with inefficient suturing. There was only 1 case of mesh erosion and vaginal protrusion. The mesh was removed.</p> <p>Postoperative complications included 10 cases of fever, 5 cases of lumbosciatica , 15 cases of detrusor overactivity, 2 cases of vaginal haematoma and 5 cases of minimal dyspareunia (2 cases persisted and 3 cases resolved spontaneously).</p>
	Rate		
Quite satisfied	83% (115/138)		
Satisfied enough	12%(16/138)		
Not satisfied	5% (7/138)		
Abbreviations used: HWS, half way system classification by Baden and Walker; LSC, laparoscopic sacrocolpopexy; SD, standard deviation.			

^aMedian follow-up 43 months (range 6-96)^bIn all patients except 25 women affected by 2nd degree vaginal vault prolapse.^cIn all patients affected by 4th degree vault prolapse and all patients affected by 3rd degree prolapse except 9.

Study 9 Campbell P (2016)

Details

Study type	Systematic review and meta-analysis
Country	UK
Recruitment period	Date limit for literature search not stated.
Study population and number	n= 1461 , from 7 trials (589 in the LSC group and 872 in the SCP group) Trials Included RCT: Freeman 2013 Case series: Coolen 2013, Nosti 2014, McDermott 2012, Hsiao 2007, Paraiso 2005 and Klauschie 2009
Age and sex	Mean age not stated, women
Patient selection criteria	Women with apical pelvic organ prolapse treated by SCP or LSC.
Technique	RCTs and observational studies that compared patients treated by LSC or SCP were included.
Follow-up	6 to 16 months
Conflict of interest/source of funding	None.

Analysis

Follow-up issues: Incomplete outcome data, which refers to long-term follow-up for objective assessment of prolapse, was graded low risk in 2 studies; Coolen 2013 collected all data prospectively, and in the study by Klauschie 2009 outcome data were unavailable in only 1 of 85 patients. The study by Freeman 2013 where 3 patients in each group were missing from the 1-year follow-up analysis, was graded “unclear,” whereas the remaining studies were all deemed high risk.

Study design issues: A manual search of reference lists of all known and included studies was conducted to identify studies not captured by electronic searches. No language restrictions were applied. The title and abstracts were screened by 2 independent reviewers. Any discrepancies were settled by discussion with the senior author. The Cochrane Collaboration tool for assessing risk of bias was used to assess methodological quality of the papers.

Study population issues: There was significant heterogeneity of the studies with respect to women undergoing concomitant surgery at the time of SCP. Women undergoing concomitant hysterectomy were included in the studies by McDermot 2012, Klauschieetal 2009, and Nosti 2014.

The study by Nosti 2014 reports outcomes of 1124 women. From these 589 were treated by SCP and 535 by either LSC (273) or RASC (262).

Other issues: The study by Freeman 2013 was also included in the paper by Maher 2016¹.

Key efficacy and safety findings

Efficacy	Safety
<p>n=1,461 (589 LSC and 872 SCP)</p> <p><u>SCP 60% (831/1377) versus LSC 40% (546/1377), 6 trials n=1377</u></p> <p>Duration of surgery (favours SCP) MD 25.25 minutes, 95% CI, 5.43–45.07 minutes $I^2=91\%$, $p=0.00001$</p> <p>Intraoperative blood loss (favours LSC) MD -106.66 mL, 95% CI, -139.59 to -73.73, $I^2=58\%$ $p=0.01$</p> <p>Hospital stay (favours LSC) MD -1.71 days, 95% CI, -2.21 to -1.22, $I^2=87\%$, $p=0.00001$</p>	<p>Bladder injury <u>SCP versus LSC, 7 trials (n=1461)</u> OR 0.99, 95% CI, 0.54–1.83</p> <p>SCP 4% (23/589) versus 2% LSC (5/273), 1 trial (n=862), $p<0.01$</p> <p>Bowel obstruction <u>SCP 5% (39/804) versus LSC 2% (8/520), 5 trials n=1324</u> OR: 2.88, 95% CI, 1.31–6.33, $I^2=$, $p=0.008$</p> <p>Mesh exposure <u>SCP versus LSC, 6 trials n=1377</u> OR: 1.45, 95% CI, 0.78–2.69, p value not reported</p> <p>Repeat surgery for prolapse, 6 trials, n=1377 SCP 1% (10/831) versus LSC 3% (15/546) OR 1.52; 95% CI 0.7–3.28</p> <p>Conversion to open SCP The conversion rate for LSC to SSC was 3% (17/589)</p>
<p>Abbreviations used: CI, confidence interval; LSC, laparoscopic sacrocolpopexy; MD, mean difference; OR, odds ratio; RCT, randomised controlled trial;; RASC, robot-assisted sacrocolpopexy; RR, risk ratio; SCP, abdominal sacrocolpopexy; SUI, stress urinary incontinence.</p>	

Study 10 Mirabile G (2016)

Details

Study type	Case report
Country	Italy
Recruitment period	2015
Study population and number	n=1
Age and sex	67 year-old woman
Patient selection criteria	Patient treated by laparoscopic sacrocolpopexy in 2010.
Technique	Report of complication
Follow-up	4 years
Conflict of interest/source of funding	Not reported.

Analysis

Follow-up issues: None.

Study design issues: None

Study population issues: None

Other issues: None.

Key efficacy and safety findings

Safety
<p>3 years after surgery</p> <p>Complaints: urinary incontinence, burning sensation, urgency.</p> <p>Findings: An ultrasonography identified a 4 cm bladder stone attached to the bladder wall at the point of the mesh erosion. An urethrocytography has shown trackers (used during the procedure) behind the stone and also a vesico-vaginal fistula.</p> <p>Patient had surgery to remove the mesh and reconstruct the vagina. The fistula persisted requiring reoperation. The patient improved and was continent at 1 year-follow-up.</p> <p>Abbreviations used: None.</p>

Efficacy

Subjective failure

In a systematic review and meta-analysis of 3,414 women from 30 randomised control trials (RCTs) comparing surgery for apical pelvic organ prolapse, 16 studies compared surgery for vaginal vault prolapse. In 3 RCTs (n=277) comparing sacrocolpopexy (SCP) had a statistically significantly lower rate of subjective failure than vaginal procedures (7% [10/139] for SCP compared with 16% [22/138] for vaginal procedures, risk ratio [RR] 2.11; 95% confidence interval [CI] 1.06 to 4.21, $I^2=0\%$). In one RCT in the same systematic review, SCP using mesh was not statistically significantly different from SCP using biological grafts in preventing subjective failure (3% [1/29] versus 10% [3/29], n=58, RR 0.33, 95% CI, 0.04 to 3.02). One RCT comparing SCP with SCP with colposuspension reported no statistically significant difference in subjective failure (31% [22/71] versus 37% [27/73] n=144, RR 1.19, 95% CI 0.75 to 1.89) at 7-year follow-up, in the same systematic review.¹

In an RCT of 215 patients comparing 111 women who had SCP alone with 104 women treated by SCP combined with urethropexy, symptomatic failure was greater (but not statistically significant different) in the urethropexy group than in the SCP alone group at 2-year follow-up, treatment difference 0.026, 95% CI -0.032 to 0.087. This remained true at 7-year follow-up with a treatment difference of 0.049, 95% CI -0.060 to 0.162.⁴

In an RCT of 100 patients that compared SCP with polypropylene mesh (n=54) and SCP with cadaveric fascia lata (n=46), there were no subjective complaints in the mesh group compared with 13% (4/29) of patients reporting vaginal bulge and 10% (3/29) of patients reporting symptoms of prolapse in the fascia lata group, within 5-year follow-up, p value not reported.⁵

In a prospective case series of 101 women treated by laparoscopic sacrocolpopexy (LSC), subjective cure rates were 98% (97/99) at the end of 12 months and 95% (81/85) at the end of 60 months, p value not reported.⁶

Objective failure

Any type of prolapse

In a systematic review of 3,414 women, 4 RCTs (n=390) reported that SCP was associated with statistically significantly less recurrent prolapse than vaginal procedures at 1 to 2-year follow-up (19% [35/189] compared with 34% [68/201], RR 1.89, 95% CI 1.33 to 2.70; $I^2=41\%$). Two RCTs (n=173) in the same systematic review reported that there was no statistically significant difference in prolapse recurrence between women in the SCP using mesh group and women treated by SCP using a biological graft (11% [10/87] versus 23% [20/86],

RR 0.49, 95% CI 0.20 to 1.25; $I^2=48\%$). Two RCTs ($n=96$) reported no statistically significant difference in the rate of prolapse recurrence between the SCP and or RASC groups (10% [5/50] versus laparoscopic sacrocolpopexy [LSC] 9% [4/46], RR 0.87; 95% CI 0.25 to 3.06) in the systematic review of 3,414 women. In 1 RCT ($n=70$) in the same systematic review, recurrent prolapse was not statistically significantly different between women treated by SCP alone and women in the SCP with colposuspension group (38% [14/37] versus 30% [15/33], RR 1.20, 95% CI 0.69 to 2.10).¹

In a systematic review of 1,176 women (from 13 comparative studies), SCP using mesh was compared with native tissue repair. Prolapse reduction success was defined as a less than stage 2 prolapse or an above the hymen measurement. For a follow-up period between 1 and 2.5 years the meta-analysis of 4 RCTs from this systematic review reported that SCP using mesh had statistically significantly better objective cure rates than native tissue vaginal repair (75% [132/177] compared with 62% [119/192], odds ratio (OR) 2.04; 95% CI 1.12 to 3.72, $I^2=31\%$).³

In a systematic review and meta-analysis of 1,488 patients from 27 studies (17 single arm and 10 comparative studies), all-compartment prolapse was 6% (66/1,029) at a minimum 2-year follow-up.³

In the RCT of 100 women comparing SCP using polypropylene mesh with SCP using cadaveric fascia lata, overall objective anatomic success was statistically significantly higher in the polypropylene group (93% [27/29]) than in the fascia lata group (62% [18/29]) at 5-year follow-up, $p=0.02$.⁵

In a case series of 165 women treated by LSC there was recurrent vault prolapse in 5% (7/138) of women, recurrent rectocele in 1% (1/138) and cystocele in 4% (5/138) of women at the end of 8-year follow-up.⁸

In the RCT of 215 women comparing 111 women treated by SCP alone with 104 women treated by SCP with urethropexy, there was no statistically significant difference in anatomic failure at 2-year follow-up (treatment difference: -0.005 , 95%CI -0.093 to 0.087) or at the end of the 7-year follow-up (treatment difference: 0.05 , 95%CI -0.161 to 0.271).⁴

In the prospective case series of 101 women treated by LSC, objective cure rate was 92% (91/99) at month 12 and 84% (57/68) after 60 months.⁶

In the prospective case series of 165 women treated by LSC, objective cure rate was 95% (131/138) within a median follow-up of 43 months.⁸

Anterior prolapse

In 2 RCTs ($n=199$) from the systematic review of 3,414 women, the rate of anterior compartment prolapse was statistically significantly less frequent in

women treated by SCP than in women treated by vaginal procedures (6% [6/102] compared with 24% [23/97], RR 4.02, 95% CI 1.71 to 9.49; $I^2=22\%$).¹

In the prospective case series of 101 women treated by LSC, recurrence of anterior wall prolapse was 6% (6/99) and 9% (6/68) at 12 and 60 months respectively.⁶

Apical prolapse

In 3 RCTs (n=275) from the systematic review of 3,414 women, apical compartment prolapse was statistically significantly less frequent in women in the SCP group than in women treated by vaginal procedures (2% [3/134] compared with 21% [29/141], RR 8.15, 95% CI 2.71 to 24.49; $I^2=0\%$).¹

In a systematic review and meta-analysis of 1,488 women, apical prolapse rate was less than 1% (2/246).³

Posterior prolapse

In 2 RCTs (n=199) posterior compartment prolapse was statistically significantly less frequent in the SCP group than in women treated by vaginal procedures (3% [3/99] compared with 12% [12/100], RR 3.43, 95% CI 1.10 to 10.66; $I^2=0\%$) in the systematic review of 3,414 women.¹

In 1 comparative study included in the systematic review of 1,176 women, there was a statistically significantly higher recurrence of posterior wall prolapse after LSC (17% [10/60]) than after sacrospinous ligament fixation (0/51, $p<0.01$).²

In the prospective case series of 101 women treated by LSC, recurrence of posterior wall prolapse was 2% (2/99) and 6% (4/68) at 12 and 60 months respectively.⁶

POP-Q measurements

Objective success was reported by some studies through specific point measurements, using the pelvic organ prolapse quantification system (POP-Q).

Point Ba (middle anterior vaginal wall)

In 1 RCT (n=108) included in the systematic review and meta-analysis of 3,414 women, women treated by SCP had statistically significantly more support on point Ba than women treated by vaginal procedures (mean difference [MD] 0.80 cm, 95% CI 0.41 to 1.19). In 1 RCT (n=58) in the same systematic review there was a statistically significantly small benefit in point Ba support favouring treatment by SCP using mesh over treatment by SCP using a biological graft (MD 0.80 cm, 95% CI 0.20 to 1.40). In 1 RCT (n=78) point Ba support was not statistically significantly different between the LSC group and the RASC group (MD 0.05 cm; 95% CI -0.31 to 0.41), in the systematic review and meta-analysis

of 3,414 women. In one RCT (n=322) in the same systematic review there was statistically significantly worse point Ba support in women treated by SCP than in women treated by SCP with colposuspension (MD -0.40 cm, 95% CI, -0.62 to -0.18).¹

Point Bp (mid-point posterior vaginal wall)

In 1 RCT (n=108) included in the systematic review and meta-analysis of 3,414 women, those treated by SCP (53) had statistically significant better support on point Bp than women treated by vaginal procedures (MD 0.77 cm, 95% CI 0.38 to 1.16). In 1 RCT (n=58) in the same systematic review, point Bp did not have statistically significantly better support in women treated by SCP using mesh than by SCP using a biological graft (MD -0.20, 95% CI, -0.51 to 0.11). In 2 RCTs (n=125) point Bp support was statistically significantly worse in the SCP or RASC group than in women treated by LSC (MD -0.40, 95% CI, -0.76 to -0.05; $I^2=0\%$), in the systematic review and meta-analysis of 3,414 women. In 1 RCT (n=322) in the same systematic review, there was statistically significantly better point Bp support in women treated by SCP than in women treated by SCP with colposuspension (MD 0.30, 95% CI 0.11 to 0.49).¹

Point C (most distal edge of cervix or vaginal cuff scar)

In 1 RCT (n=108) included in the systematic review and meta-analysis of 3,414 women, women treated by SCP had statistically significantly more support on point C than women treated by vaginal procedures (MD 0.50 cm, 95% CI 0.11 to 0.88). In 1 RCT (n=58) in the same systematic review, point C did not have statistically significantly better support with treatment by SCP using mesh than by SCP using a biological graft (MD 0.31 cm, 95% CI -0.41 to 1.03). In 3 RCTs (n=197) point C support was not statistically significantly different between the SCP or RASC group and the LSC group (MD 0.15 cm, 95% CI -0.52 to 0.83, $I^2=0\%$), in the systematic review and meta-analysis of 3,414 women. One RCT (n=322) of the same systematic review reported no difference in point C support in women treated by SCP than in women treated by SCP with colposuspension (MD 0.20 cm, 95% CI -0.11 to 0.51).¹

Total vaginal length

In 1 RCT (n=108) included in the systematic review and meta-analysis of 3,414 women, women treated by SCP had a statistically significantly longer vagina than women treated by vaginal procedures (MD -0.89 cm, 95%CI -1.29 to -0.50). In 1 RCT (n=58) in the same systematic review, vaginal length was not statistically significantly different in women treated by SCP using mesh compared with women treated by SCP using a biological graft (MD -0.10 cm, 95% CI -0.69 to 0.49).¹

In the RCT of 100 patients comparing SCP with polypropylene mesh (29/54) to SCP with cadaveric fascia lata (20/46) there was no statistically significant difference for any of the POP-Q measurements at the end of 5-year follow-up.⁵

Patient satisfaction and quality of life

In 1 RCT (n=110) in the systematic review and meta-analysis of 3,414 women, there was no statistically significant difference in quality of life measured by the pelvic organ prolapse/urinary incontinence sexual questionnaire (PISQ) between the SCP group and the vaginal procedures group at 4-year follow-up (MD -1.2, 95% CI -4.35 to 1.95). The same RCT reported no statistically significant difference in quality of life measured by the prolapse quality of life questionnaire (P-QOL, 0 to 100, 0=good quality of life, 100=poor quality of life) between the SCP group and the vaginal procedures group at 4-year follow-up (MD 22.70, 95% CI -7.53 to 52.93). In 1 RCT (n=47) from the same systematic review of 3,414 women there was no statistically significant difference in quality of life measured by P-QOL in women treated by SCP compared with women treated by LSC (MD 0.70, 95% CI, -19.14 to 20.54). In 1 RCT (n=115) there was no statistically significant difference in quality of life measured by the pelvic floor impact questionnaire (PFIQ-7) between the SCP using mesh and the SCP using biological graft groups (MD -7.00, 95% CI -29.48 to 15.48), in the systematic review of 3,414 women. In 1 RCT (n=78) in the same systematic review, there was no statistically significant difference in quality of life measured by PFIQ-7 between women treated by LSC and women treated by RASC (MD 21.00; 95% CI -46.76 to 88.76). In 1 RCT (n=110) a statistically significantly better quality of life, measured by the pelvic floor distress inventory (PFDI-20), was reported in women treated by SCP than in women treated by vaginal procedures (MD 7.90; 95% CI 0.70 to 15.10), in the systematic review of 3,414 women. In 1 RCT (n=115) in the same systematic review there was no statistically significant difference in quality of life measured by PFDI-20 between the SCP using mesh group and the SCP using biological graft group (MD -6.00, 95% CI, -25.75 to 13.75).¹

In the prospective case series of 101 women treated by LSC, the quality-of-life score improved from 5.6 at baseline to 9.1 at 12 months and 8.3 at 60 months (measured on a visual analogue scale between 1 and 10).⁶

In a prospective case series of 70 women treated by robot-assisted sacrocolpopexy (RASC), 55% (22/40) would recommend the procedure to a relative or friend, 25% (10/40) would probably recommend the procedure and overall satisfaction was 10 (0=not at all success, 10=very successful) at the median follow-up time of 90 months. The average symptomatic improvement was 9 (0=much worse, 10=much better). The median scores for the pelvic floor distress inventory (PFDI-20) (score 25–100) were: POP distress inventory, 29.2 (IQR 25–37.5), colorectal-anal distress inventory-8, 40.6 (IQR 28.1–47.7) and urinary distress inventory-6, 45 (IQR 35–60). The median total score for the

pelvic floor impact questionnaire short form 7 (PFIQ-7) (score 0–300) was 0 (0–28.6).⁷

In the prospective case series of 165 women treated by LSC, 83% (115/138) of women were 'quite satisfied', 12% (16/138) were 'satisfied enough' and 5% (7/138) were 'not satisfied'.⁸

Stress urinary incontinence

In the systematic review and meta-analysis of 3,414 women, 3 RCTs (n=263) reported that stress urinary incontinence (SUI) was statistically significantly lower in the SCP group than in the vaginal procedures group (17% [21/127] versus 31% [42/136], RR 1.86, 95% CI 1.17 to 2.94; $I^2=0$). In 1 RCT (n=73) from the same systematic review there was a small statistically significant difference in SUI favouring the LSC group compared with the SCP or RASC groups (5% [2/38] versus LSC 9% [3/35], RR 1.63 95% CI 0.29 to 9.18). In 3 RCTs (n=295) there was no statistically significant difference in SUI between women treated by SCP alone and by SCP with colposuspension at 4 to 7 years follow-up (random-effects RR 1.13, 95% CI 0.63 to 2.04; 3 RCTs $I^2=70\%$), in the systematic review of 3,414 women.¹ In the RCT of 215 women the rate of SUI was statistically significantly higher in the SCP only group when compared with the SCP with urethropexy group, treatment difference: -0.175; 95% CI -0.296 to -0.043, at 2-year follow-up. Treatment difference loses statistical significance in overall urinary incontinence: -0.082; 95%CI -0.203 to 0.041. The trend remains at the end of 7-year follow-up, with rates of SUI being statistically significantly lower in the urethropexy group, treatment difference: -0.154; 95%CI -0.266 to -0.037. The treatment difference for overall causes of urinary incontinence was not statistically significant, treatment difference: -0.064; 95% CI -0.161 to 0.032.⁴

In the prospective case series of 165 women a number of urinary symptoms improved at the end of 43-month follow-up: nycturia complaints reduced from 17% (24/138) to 4% (4/138), dysuria reduced from 9% (13/138) to 3% (4/138), mixed incontinence decreased from 17% (23/138) to 14% (20/138), pollakiuria reduced from 13% (18/138) to 7% (10/138), voiding dysfunction decreased from 16% (22/138) to 7% (9/138) and the incidence of recurrent urinary tract infection also decreased from 16% (22/138) to 5% (7/138). Some symptoms had worsened at the end of the 43 months. These included stress incontinence that was reported by 4% (5/138) of the women and increased to 7% (11/138), urge incontinence that increased from 11% (15/138) to 18% (25/138).⁸

Improvement in bowel symptoms

The prospective case series of 165 women reported that constipation rates increased from 7% (10/138) before surgery to 13% (18/138) at the end of follow-up, and obstructed defaecation increased from 1% (2/138) to 6% (8/138). Urgency was not reported by any women before surgery and it was reported in 2% (3/138) of women at the end of 43 months. The incidence of pelvic pressure

symptoms reduced from 67% (92/138) to 9% (12/138) at the end of follow-up. Similarly, the incidence of false urge to defaecate reduced from 51% (70/138) of women at baseline to 5% (7/138) at 43 months.⁸

Dyspareunia

In the systematic review and meta-analysis of 5,954 women from 56 RCTs, in 3 RCTs reduction in postoperative dyspareunia was greater in the SCP group than in the VSC group (16% [7/45] for VSC compared with 36% [22/61] for SCP, RR 0.39; 95%CI 0.18 to 0.86).¹

The dyspareunia rate was statistically significantly lower in women treated by SCP using mesh (5% [23/445]) than in women treated by native tissue repair (12% [46/384], OR 0.42; 95% CI 0.25 to 0.72) from the analysis of 5 comparative studies reported in the systematic review and meta-analysis of 1,176 women. The rate of dyspareunia was similar for SCP using mesh (12% [371/2,986]) and native tissue repair (9% [200/2,180]; $p=0.48$) in the analysis of non-comparative studies in the same systematic review.²

Improvement in symptoms

In the prospective case series of 165 women, a vaginal lump was reported by 83% (115/138) of women before treatment and by 7% (10/138) at 43-month follow-up. Similarly, lower abdominal pain was present in 69% (95/138) of the women before the procedure and in 4% (6/138) at 43 months.⁸

Safety

Death and admission to ICU

Incidence of death was not statistically significantly different in women treated by abdominal sacrocolpopexy (SCP) using mesh (0/503) compared with women treated using native tissue (less than 1% [1/582]; OR: 0.14; 95% CI 0.003 to 6.97) in the analysis of comparative studies reported in the systematic review and meta-analysis of 1,176 women. Postoperative admission to intensive care was not statistically significantly different between the SCP using mesh group (1% [3/561]) compared with the native tissue repair group (0/506; OR 4.64; 95% CI 0.42 to 50.6) in the analysis of comparative studies in the same systematic review.²

Incidence of death was not statistically significantly different in women treated by SCP using mesh (<1% [6/3343]) compared with women treated using native tissue (<1% [12/4105], $p=0.61$) in the analysis of non-comparative studies reported in the systematic review and meta-analysis of 1,176 women.

Postoperative admission to intensive care was not statistically significantly different in the SCP using mesh group (6% [281/4233]) compared with the native tissue repair group (1% [27/3532], $p=0.11$) in the analysis of non-comparative studies in the same systematic review.²

Deep vein thrombosis and pulmonary embolism

Deep vein thrombosis or pulmonary embolism was not statistically significantly different between the SCP using mesh group (less than 1% [2/569]) and the native tissue repair group (less than 1% [1/599], OR 1.36; 95% CI 0.14 to 13.7) in women included in comparative studies from the same analysis. Deep vein thrombosis or pulmonary embolism was statistically significantly higher in the SCP using mesh group (1% [46/4,579]) than in the native tissue repair group (less than 1% [8/4,114], $p=0.03$) in women included in non-comparative studies from the same analysis.²

Mesh erosion

Mesh exposure risk was not statistically significantly different in women treated by SCP (3% [8/283]) compared with vaginal procedures (4% [9/291] RR: 1.13; 95% CI 0.47 to 2.69; $I^2=28\%$), in 6 RCTs ($n=574$) reported in the systematic review and meta-analysis of 3,414 women. Mesh exposure risk was not statistically significantly different in the SCP using mesh group 3% (3/86) compared with the SCP using a biological graft group (1% [1/87], RR: 2.35, 95% CI 0.36 to 15.40, $I^2=0\%$) at 1 to 5 years follow-up in 2 RCTs ($n=173$) reported in the systematic review and meta-analysis of 3,414 women. Mesh exposure risk was not statistically significantly different in women treated by SCP or RASC (2% [2/96]) compared with LSC (0/90, RR: 0.22, 95% CI 0.01 to 4.40, $I^2=0\%$) in 3 RCTs ($n=186$) in the same systematic review.¹

Mesh or suture complications were statistically significantly more frequent in women treated by SCP using mesh (4% [28/650]) than in women who had native tissue repair (1% [6/537], OR 3.26; 95% CI 1.62 to 6.56) in an analysis of comparative studies in the systematic review of 1,176 women. Mesh or suture complications were statistically significantly more frequent in women treated by SCP using mesh (4% [348/7,831]) than in women treated by native tissue repair (less than 1% [13/1,169], $p<0.001$) in the analysis of 40 SCP compared with 11 native tissue repair non-comparative studies.²

Mesh erosion was not statistically significantly different between robot-assisted sacrocolpopexy (RASC) and laparoscopic sacrocolpopexy (LSC; OR 1.82; 95%CI 0.51 to 6.45 [$n=438$], $I^2=0\%$, $p=0.86$) in the systematic review of 1,488 women. Mesh erosion was statistically significantly lower in women treated by RASC with supracervical hysterectomy (0%) than in women treated by RASC after total hysterectomy (14%, $p=0.008$) in 1 comparative study included in the same systematic review.³

Mesh erosion occurred with all types of mesh with an overall probability of 11% (95% CI 7% to 16%) at 6-year follow-up when the right censoring time was the last clinic visit in the RCT ($n=215$) comparing 104 patients who had SCP combined with urethropexy with 111 women who had SCP alone. The mesh erosion rate was 10% (95% CI 7% to 15%) when the right censoring time was either a clinic visit or a last telephone interview.⁴

Graft erosion occurred with the same frequency (1 woman in each group) in the group of women treated by SCP with polypropylene mesh compared with SCP with cadaveric fascia lata at 5-year follow-up in the RCT of 100 patients. The polypropylene mesh erosion occurred at the posterior wall and eroded to the rectum. The woman needed bowel resection and a colostomy. The woman was lost to follow-up. The woman with fascia lata graft erosion was lost to follow-up between years 1 and 5. At the year 5 visit the fascial erosion persisted and the patient presented with post-coital spotting, dyspareunia, vaginal discharge and odour. At 5-year follow-up there was 1 additional erosion in the polypropylene mesh group. The woman had a 2×3 cm apical mesh erosion. Laparoscopic vaginal removal of the mesh was needed. Necrotising fasciitis developed post-operatively at the umbilical port site. The patient had a long stay in hospital but fully recovered.⁵

Mesh erosion was reported in 1% (1/99) of women at 12 months and 3% (2/85) at 60 months in women treated by LSC in the prospective case series of 101 women.⁶

Mesh exposure risk was not statistically significantly different in women treated by SCP compared with LSC (OR: 1.45, 95% CI 0.78 to 2.69) in 6 trials (n=1,377) in the systematic review and meta-analysis of 1,461 women.⁹

Repeated surgery

Pooled estimates

Reoperation rates were similar for women treated by SCP or sacrospinous ligament fixation (SSLF, 13% [6/46] versus 16% [7/43], p=0.67) in an RCT (reported in the systematic review of 1,176 women) with follow-up of 6 to 66 months. Pooled reoperation rates were 7% (46/615) for SCP and 10% (51/511) for native tissue repair (OR 0.76, 95% CI 0.28 to 1.09) in 7 comparative studies from the same systematic review and meta-analysis. Pooled reoperation rates in non-comparative studies were 5% (367/7,218) for SCP and 3% (114/3,872) for native tissue repair, (p=0.28) in the systematic review of 1,176 women.²

The reoperation rate was 3% (23/687) in women treated by RASC in the systematic review and meta-analysis of 1,488 patients from 27 studies. A feeling of traction needing reoperation was reported in less than 1% (2/1,118) of the women treated by RASC in the same systematic review.³

Additional surgery was needed in 16% (36/215) of the women included in the RCT (n=215) of 104 women treated by SCP combined with urethropexy compared with 111 women treated by SCP alone, at 7-year follow-up. Causes of reoperation were 11 recurrent POP, 14 for stress urinary incontinence and 11 mesh complications.⁴

The reoperation rate was similar in women treated by SCP with polypropylene mesh (3% [1/29]) compared with SCP with cadaveric fascia lata (3% [1/29]) in the RCT of 100 patients.⁵

Reoperation rates in women treated by RASC were 2%, 5% and 10% at years 1, 3 and 6 respectively in the prospective case series of 70 women.⁷

Repeated surgery for mesh exposure

Repeated surgery for mesh exposure was not statistically significantly different in women treated by SCP compared with vaginal procedures, (RR 1.14; 95% CI 0.35 to 3.64; $I^2=48\%$) in 5 RCTs (n=497) in the systematic review and meta-analysis of 3,414 women. Repeated surgery for mesh exposure was not statistically significantly different in women treated by SCP using mesh (2% [2/86]) compared with SCP using a biological graft (1% [1/87], RR: 2.00, 95% CI 0.19 to 20.86, $I^2=0\%$) in 2 RCTs (n=173) in the same systematic review and meta-analysis.¹

Repeated surgery for prolapse

Repeated surgery for prolapse was statistically significantly less frequent in women treated by SCP (6% [11/187]) than with vaginal procedures (14% [28/196], RR 2.28, 95% CI 1.20 to 4.32, $I^2=0\%$) in 4 RCTs (n=383) in the systematic review and meta-analysis of 3,414 women. Repeated surgery for prolapse was not statistically significantly different at 1 to 5 years follow-up between women treated by SCP using mesh (1% [1/87]) and by SCP using a biological graft (1% [1/86], RR: 1.00, 95% CI, 0.07 to 15.24, $I^2=0\%$) in 2 RCTs (n=173) in the same systematic review and meta-analysis. Repeated surgery for prolapse was not statistically significantly different between women treated by SCP (8% [2/24]) and by LSC (9% [2/23], RR 1.04, 95% CI, 0.16 to 6.80) in 1 RCT (n=47) in the systematic review and meta-analysis of 3,414 women. Repeated surgery for prolapse was not statistically significantly different between the SCP alone group (5% [7/128]) and the SCP with colposuspension group (4% [5/128], RR: 0.71, 95% CI 0.24 to 2.15, $I^2=0\%$) in 3 RCTs (n=256) in the same systematic review and meta-analysis.¹

Repeated surgery for prolapse was not statistically significantly different between the SCP (1% [10/831]) and LSC groups (3% [15/546], OR: 1.52; 95% CI 0.7 to 3.28) in 7 trials (n=1,461) in the systematic review and meta-analysis of 1,461 women.⁹

Repeated surgery for SUI

Repeated surgery for SUI was not statistically significantly different between women treated by SCP (3% [6/189]) and by vaginal procedures (6% [12/206],

RR: 1.87, 95% CI 0.72 to 4.86, $I^2=0\%$) in 4 RCTs (n=395) in the systematic review and meta-analysis of 3,414 women. Repeated surgery for SUI was not statistically significantly different between the SCP using mesh group (3% [1/29]) and the SCP using a biological graft group (0/29, RR 3.00, 95% CI 0.13 to 70.74) in 1 RCT (n=58) in the same systematic review and meta-analysis. Repeated surgery for SUI was not statistically significantly different between women treated by SCP alone (5% [5/92]) and by SCP with colposuspension (8% [7/91], RR 1.42, 95% CI 0.47 to 4.30) in 1 RCT (n=183) in the systematic review and meta-analysis of 3,414 women.¹

Reoperation for SUI in women treated by LSC was reported in 15% (15/99) and 19% (16/85) of patients at 12 and 60 months respectively in the prospective case series of 101 women.⁶

Incidence of damage to surrounding structures

Nerve injury incidence was not statistically significantly different in patients treated by SCP using mesh (1% [4/514]) compared with native tissue repair (1% [7/743] OR: 0.61 [0.18 to 2.05]) in 5 comparative studies included in the systematic review and meta-analysis of 1,176 women. The incidence of nerve injury was not statistically significantly different in women treated by SCP (4% [96/2,601]) compared with native tissue repair (5% [147/2,813]) in the analysis of non-comparative studies included in the same systematic review.²

The vaginotomy rate in patients treated by RASC was 1% (14/1,488) in the systematic review and meta-analysis of 1,488 patients from 27 studies. Port site nerve entrapment happened in 1% (1/1,118) of patients in the same systematic review.³

Injury to bladder or urethra

Bladder injury was not statistically significantly different between women treated by SCP (2% [4/244]) and by vaginal procedures (1% [2/267], RR: 0.57, 95% CI 0.14 to 2.36; $I^2=0\%$) in 5 RCTs (n=511) in the systematic review and meta-analysis of 3,414 women. Similarly, bladder erosion was not statistically significantly different between the SCP using mesh (less than 1% [1/111]) and SCP using biological graft groups (0/113 groups, RR: 2.51, 95% CI 0.10 to 60.13, $I^2=0\%$) in 2 RCTs (n=224) in the systematic review and meta-analysis of 3,414 women. Bladder injury was not statistically significantly different in women treated by SCP or RASC (2% [2/102]) in comparison to LSC (4% [4/97], RR 1.75, 95% CI, 0.43 to 7.14, $I^2=0\%$) in 3 RCTs (n=199) in the same systematic review and meta-analysis.¹

Urinary tract injury was not statistically significantly different between patients treated by SCP using mesh (2% [20/1,068]) and by native tissue repair (1% [9/1,108], OR: 1.68, 95% CI 0.79 to 3.55) in 8 comparative studies from the systematic review and meta-analysis of 1,176 women. Urinary tract injury was statistically significantly higher in women treated by SCP using mesh (2%

[113/6,894]) compared with native tissue repair (1% [46/5,111], $p<0.05$) in the analysis of non-comparative studies from the same systematic review and meta-analysis.²

Bladder injury in patients treated by RASC was 2% (26/1,488) in the systematic review and meta-analysis of 1,488 patients from 27 studies. Ureteral injury incidence was less than 1% (1/1,488) in patients from the same systematic review.³

Incidental bladder incision happened in 2 patients treated by LSC in the prospective case series of 101 women. Mesh protrusion happened in the same patient: protruded mesh was removed by laparoscopy with partial excision of the anterior mesh and reconstruction of the bladder.⁶

Bladder injury was not statistically significantly different in the SCP (60% [872/1,461]) and LSC groups (40% [589/1,461], OR 0.99, 95% CI 0.54 to 1.83) in 7 trials (n=1,461) in the systematic review and meta-analysis of 1,461 women.⁹

Bowel injury

Bowel injury was not statistically significantly different in women treated by SCP (1% [2/143]) compared with vaginal procedures (less than 1% [1/163], RR: 0.63, 95% CI 0.12 to 3.23, $I^2=0\%$) in 3 RCTs (n=306) in the systematic review and meta-analysis of 3,414 women. Similarly, bowel injury risk was not statistically significantly different in women treated by SCP or RASC (4% [2/56]) compared with LSC (0/52, RR 0.36, 95% CI 0.04 to 3.32, $I^2=0\%$) in 2 RCTs (n=108) in the systematic review and meta-analysis of 3,414 women.¹

Bowel injury in women treated by SCP using mesh was not statistically significantly different (1% [8/1,219]) compared with native tissue repair (1% [10/1,574], OR 0.91; 95% CI 0.35 to 2.37) in the analysis of comparative studies from the systematic review and meta-analysis of 1,176 patients. Bowel injury rate was not statistically significantly different in women treated by SCP using mesh (1% [37/6,642]) compared with women treated by native tissue repair (1% [47/5,744], $p=0.33$), in the analysis of non-comparative studies from the same systematic review.²

Bowel injury in women treated by RASC was less than 1% (4/1,488) in the systematic review and meta-analysis of 1,488 patients from 27 studies.³

De novo urinary incontinence

De novo urge incontinence risk was not statistically significantly different in women treated by SCP (21% [6/29]) compared with vaginal procedures (33% [11/33], RR 1.61, 95% CI 0.68 to 3.81) in 1 RCT (n=62) in the systematic review and meta-analysis of 3,414 women. De novo urinary voiding dysfunction was not statistically significantly different in women treated by SCP (3% [1/38]) compared with vaginal procedures (3% [1/37], RR 1.03, 95% CI 0.07 to 15.82) in 1 RCT (n=75) in the systematic review and meta-analysis of 3,414 women.¹

De novo SUI rate in women treated by LSC was 24% (24/99) and 38% (32/85) at 12 and 60 months respectively in the prospective case series of 101 women. Postoperative voiding disorders occurred in 8% (8/99) and 13% (11/85) of women at 12 and 60 months respectively in the same patient group. De novo urge incontinence occurred in 2% (2/99) women at 12 months and in 8% (7/85) at 60 months.⁶

The detrusor muscle overactivity rate was 9% (15/165) in the case series of 165 women.⁸

De novo dyspareunia

De novo dyspareunia in women treated by LSC was 2% (1/47) and 24% (10/41) at 12 and 60 months respectively in the prospective case series of 101 women.⁶

Minimal dyspareunia in women treated by LSC was 3% (5/165) in the case series of 165 women. It was reported that this persisted in 2 women and in 3 women it resolved spontaneously).⁸

De novo prolapse

De novo rectocele in patients treated by LSC was 12% (16/138) at the end of 8 years follow-up in the case series of 165 women. Cystocele rate was 8% (11/183) in women from the same study.⁸

Infection

Infection rates were not statistically significantly different between women treated by SCP using mesh (3% [17/676]) and women treated by native tissue repair (1% [9/617], OR 2.01; 95% CI 0.91 to 4.45) in the analysis of comparative studies reported in the systematic review and meta-analysis of 1,176 women. Infection rates were not statistically significantly different between women treated by mesh SCP (2% [114/5,519]) and by native tissue repair (12% [558/4,743], $p=0.6$) in the analysis of non-comparative studies for the same systematic review.²

Abscess formation in patients treated by RASC was less than 1% (3/1,118) in the systematic review and meta-analysis of 1,488 patients from 27 studies. Peritonitis caused by bowel injury happened in less than 1% (2/1,118) patients in the same analysis.³

Fever in women treated by LSC was 6% (10/165) in the case series of 165 women.⁸

Bleeding

Bleeding rates were not statistically significantly different between women treated by SCP using mesh (3% [43/1,317]) and by native tissue repair (2% [37/1,863], OR 1.00; 95% CI 0.63 to 1.59) in the comparative studies reported in the systematic review and meta-analysis of 1,176 women. Bleeding rates were statistically significantly lower in women treated by mesh SCP (2% [128/6,555]) than by native tissue repair (5% [367/7,044], $p=0.05$) in the analysis of non-comparative studies for the same systematic review.²

Vaginal haematoma in patients treated by LSC was reported in 1% (2/165) of patients in the case series of 165 women.⁸

Intraoperative bleeding was statistically significantly higher in women treated by SCP (60% [831/1377]) compared with LSC (40% [(546/1377], MD: -106.66 mL, 95% CI -139.59 to -73.73, $I^2=58\%$) in the systematic review and meta-analysis of 1,461 women.⁹

Blood transfusion

Blood transfusion was not statistically significantly different in women treated by SCP (3% [3/91]) compared with vaginal procedures (0/97, RR: 0.26, 95% CI 0.04 to 1.57, $I^2 = 0\%$) in 3 RCTs ($n=277$) in the systematic review and meta-analysis of 3,414 women. Blood transfusion was not statistically significantly different in women treated by SCP (9% [3/32]) compared with SCP with colposuspension (9% [3/34], RR: 0.94, 95% CI 0.20 to 4.33) in 1 RCT ($n=66$) in the same systematic review.¹

Bowel obstruction

Bowel obstruction was not statistically significantly different between the SCP (60% [872/1461]) and LSC groups (40% [89/1461], OR 0.99, 95% CI 0.54 to 1.83) in 7 trials ($n=1,461$) in the systematic review and meta-analysis of 1,461 women.⁹

Ileus or small bowel obstruction was statistically significantly higher in patients treated by SCP using mesh (2% [16/814]) than in patients treated by native tissue repair (less than 1% [2/780], OR 9.45; 95% CI 3.39 to 26.4) in the analysis of comparative studies reported in the systematic review and meta-analysis of 1,176 women. Ileus or small bowel obstruction was also statistically significantly higher in women treated by mesh SCP (3% [137/4,168]) than by native tissue repair (less than 1% [3/1,449], $p<0.01$) in the analysis of non-comparative studies for the same systematic review.²

Bowel obstruction in women treated by RASC was less than 1% (5/1,118) in the systematic review and meta-analysis of 1,488 patients from 27 studies.³

Postoperative constipation in women treated by LSC was 1% (1/99) and 5% (4/85) at 12 and 60 months respectively, in the prospective case series of 101 women.⁶

Pain

Lumbosciatica pain was reported in 3% (5/165) of women treated by LSC in the case series of 165 women.⁸

Conversion to open surgery

The conversion rate of LSC to SSC was 3% (17/589) in the systematic review and meta-analysis of 1,461 women.⁹

Corneal abrasion

Corneal abrasion was not statistically significantly different in women treated by LSC (0/33) compared with RASC (3% [1/33], $p=0.99$) in 1 RCT ($n=66$) in the systematic review and meta-analysis of 3,414 women.¹

Other

Port site hernia in women treated by RASC was less than 1% (6/1,118) in the systematic review and meta-analysis of 1,488 patients from 27 studies. Vaginal cuff dehiscence occurred in 1 patient in the same systematic review. The difference in intraoperative complication rates was not statistically significant when comparing RASC with LSC (OR 1.05; 95% CI 0.52 to 2.12 [$n=443$], $I^2=0\%$, $p=0.94$) in a meta-analysis of this systematic review. Surgical conversion to open surgery was also not statistically significantly different when comparing the RASC and LSC treatment groups (OR: 0.89; 95% CI 0.25 to 3.19 [$n=443$], $I^2=0\%$, $p=0.72$). The incidence of all postoperative complications was not statistically significant when comparing RASC with LSC (OR 1.85; 95% CI 0.96 to 3.75 [$n=350$], $I^2=37\%$, $p=0.18$) and this was also true for severe postoperative complications (grade 3 or higher; OR 0.56; 95%CI 0.36 to 2.83 [$n=430$], $I^2=24\%$, $p=0.73$).³

Related bladder stone and vesico-vaginal fistula requiring repeated surgical reconstruction were reported in the case report of 1 patient treated by LSC.¹⁰

Validity and generalisability of the studies

- There is some heterogeneity between the patients included in the analysis. This can be partly explained by the number of different sacrocolpopexy procedures. Multiple approaches such as open abdominal, laparoscopic and robotic can be used. The composition and weight of the mesh necessary for the treatment also varied between studies.
- There are different types of surgery that can be done with sacrocolpopexy. So it is not always possible to distinguish the immediate and long-term safety and efficacy outcomes of sacrocolpopexy alone.
- There are 3 systematic reviews and a meta-analysis included in the analysis. These cover different approaches to sacrocolpopexy, different types of mesh and a range of procedures done at the same time.¹⁻³ The remaining papers in table 2 include 2 randomised control trials with a maximum follow-up of 7 years⁴ and 3 prospective case series with follow-ups between 4 and 6 years⁵⁻⁸.

Existing assessments of this procedure

In December 2015, the Scientific Committee on Emerging and Newly Identified Health Risks (SCENIHR) published 'The safety of surgical meshes used in urogynaecological surgery'. It stated: The SCENIHR considers 3 factors as being important when assessing the risks associated with mesh application: the overall surface area of material used, the product design and the properties of the material used. In addition, the available evidence suggests a higher morbidity in treating female pelvic organ prolapse (POP) than for treating stress urinary incontinence (SUI), as the former uses a much larger amount of mesh.¹¹

The body of evidence suggests that when assessing the health risks of synthetic meshes, there is a need to clearly separate the smaller risks associated with

stress urinary incontinence sling surgery from those of pelvic organ prolapse mesh surgery.¹¹

Based on the currently marketed products, assessment of the risks reported indicates that polypropylene type 1 meshes are the most appropriate synthetic meshes for vaginal use and polypropylene type 1 and polyester type 3 for insertion via the abdominal route. However, there is a need for further improvement in the composition and design of synthetic meshes, in particular for female pelvic organ prolapse surgery.¹¹

SCENIHR's recommendations include:

- Material properties, product design, overall mesh size, route of implantation, patient characteristics, associated procedures (e.g. hysterectomy) and surgeon's experience are aspects influencing the clinical outcome following mesh implantation. Such aspects are to be considered when choosing appropriate therapy.
- For all procedures, the amount of mesh should be limited where possible.
- The implantation of any mesh for the treatment of POP via the vaginal route should be only considered in complex cases in particular after failed primary repair surgery.
- A certification system for surgeons should be introduced based on existing international guidelines and established in cooperation with the relevant European Surgical Associations.¹¹

A mesh working group interim report was published in December 2015 by NHS England. Its recommendations included: reviewing the current NICE guidance and creating new guidance, raising awareness among GPs of complications and how to address them, improving rates of reporting of adverse events to the Medicines and Healthcare products Regulatory Agency (MHRA), and submissions to the British Society of Urogynaecology (BSUG) and British

Association of Urological Surgeons (BAUS) databases, improving HES coding, raising awareness among patients of their option to use MHRA reporting procedures for adverse incidents, and developing information leaflets on mesh implant procedures for both SUI and POP, which provide consistent and understandable information to be used in the consenting process.¹²

In February 2016 the Royal College of Obstetricians and Gynaecologists (RCOG) published an addendum updating its guidance on the management of post-hysterectomy vaginal vault prolapse. The document states laparoscopic sacrocolpopexy is as effective as abdominal sacrocolpopexy in the management of vaginal vault prolapse.¹³

Related NICE guidance

Below is a list of NICE guidance related to this procedure. Appendix B gives details of the recommendations made in each piece of guidance listed.

- Sacrocolpopexy with hysterectomy using mesh to repair uterine prolapse. NICE interventional procedure guidance IPG577 (2017). Available from <https://www.nice.org.uk/guidance/IPG577>
- Single-incision short sling mesh insertion for stress urinary incontinence in women. NICE interventional procedure guidance IPG566 (2016). Available from <https://www.nice.org.uk/guidance/IPG566>
- Insertion of mesh uterine suspension sling (including sacrohysteropexy) for uterine prolapse repair. NICE interventional procedure guidance IPG282 (2009). Available from <https://www.nice.org.uk/guidance/IPG282>
- Infracoccygeal sacropexy using mesh for vaginal vault prolapse repair. NICE interventional procedure guidance IPG281 (2009). Available from <https://www.nice.org.uk/guidance/IPG281>
- Infracoccygeal sacropexy using mesh for uterine prolapse repair. NICE interventional procedure guidance IPG280 (2009). Available from <https://www.nice.org.uk/guidance/IPG280>

- Surgical repair of vaginal wall prolapse using mesh. NICE interventional procedure guidance IPG267 (2008). Available from <https://www.nice.org.uk/guidance/IPG267>

NICE guidelines

- Urinary incontinence in women NICE guideline CG171 (2013). Available from <https://www.nice.org.uk/guidance/cg171>

Specialist advisers' opinions

Specialist advice was sought from consultants who have been nominated or ratified by their Specialist Society or Royal College. The advice received is their individual opinion and is not intended to represent the view of the society. The advice provided by Specialist Advisers, in the form of the completed questionnaires, is normally published in full on the NICE website during public consultation, except in circumstances but not limited to, where comments are considered voluminous, or publication would be unlawful or inappropriate. Two Specialist Adviser Questionnaires for sacrocolpopexy using mesh for vaginal wall prolapse repair were submitted and can be found on the [NICE website](#).

Patient commentators' opinions

Fourteen commentaries from patients who had experience of this procedure were received, which were discussed by the committee.

Company engagement

A structured information request was sent to 3 companies who manufacture a potentially relevant device for use in this procedure. NICE received 1 completed submission. This was considered by the IP team and any relevant points have been taken into consideration when preparing this overview.

Issues for consideration by IPAC

- Two of the systematic reviews and meta-analyses included in table 2 include patients that had other procedures such as hysteropexy, cervicopexy and hysterectomy concomitantly to sacrocolpopexy. Subgroup analysis was not always available.

- NCT02676973: Apical Suspension Repair for Vault Prolapse in a Three-Arm Randomized Trial Design (ASPIRe). Study type: multi-center RCT; population: 363 women adult treated by sacral colpopexy compared with apical transvaginal mesh post-hysterectomy; location: USA; study start date: March 2016; estimated completion date: February 202; estimated study completion date=April 202. Responsible party: National Institute of Child Health and Human Development Pelvic Floor Disorders Network.

References

1. Maher C, Feiner B, Baessler K et al. (2016) Surgery for women with apical vaginal prolapse. The Cochrane database of systematic reviews 10, CD012376.
2. Siddiqui NY, Grimes CL, Casiano ER et al. (2015) Mesh sacrocolpopexy compared with native tissue vaginal repair: A systematic review and Meta-analysis. *Obstetrical and Gynecological Survey* 70:250-251.
3. Serati M, Bogani G, Sorice P et al. (2014) Robot-assisted sacrocolpopexy for pelvic organ prolapse: a systematic review and meta-analysis of comparative studies. *European Urology* 66:303-18.
4. Nygaard I, Brubaker L, Zyczynski HM et al. (2013) Long-term outcomes following abdominal sacrocolpopexy for pelvic organ prolapse. *JAMA* 309:2016-2024.
5. Tate S B, Blackwell L, Lorenz DJ et al. (2011) Randomized trial of fascia lata and polypropylene mesh for abdominal sacrocolpopexy: 5-year follow-up. *International Urogynecology Journal* 22:137-143.
6. Sarlos D, Kots L, Ryu G et al. (2014) Long-term follow-up of laparoscopic sacrocolpopexy. *International Urogynecology Journal* 25:1207-1212.
7. Linder BJ, Chow GK, and Elliott DS (2015) Long-term quality of life outcomes and retreatment rates after robotic sacrocolpopexy. *International Journal of Urology* 22:1155-1158.
8. Granese R, Candiani M, Perino A et al. (2009) Laparoscopic sacrocolpopexy in the treatment of vaginal vault prolapse: 8 years experience. *European Journal of Obstetrics, Gynecology, and & Reproductive Biology* 146:227-231.
9. Campbell P, Cloney L, Jha Swati (2016) Abdominal Versus Laparoscopic Sacrocolpopexy: A Systematic Review and Meta-analysis. *Obstetrical & gynecological survey* 71(7), 435-42.
10. Mirabile G, Rossetti A, Gentile BC et al. (2016) Vesico-vaginal fistula and bladder stone caused by a protruding spiral tacker 4 years after a laparoscopic sacrocolpopexy: Case report. *Archivio italiano di urologia, and andrologia: organo ufficiale [di] Societa italiana di ecografia urologica e nefrologica / Associazione ricerche in urologia* 88(1), 64-5.
11. SCENIHR (Scientific Committee on Emerging and Newly Identified Health Risks), The safety of surgical meshes used in urogynecological surgery, 3 December 2015.

12. NHS England, Acute Care Policy and Strategy Unit. Mesh working group interim report. Published on 3 December 2015. <https://www.england.nhs.uk/wp-content/uploads/2015/12/mesh-wg-interim-rep.pdf>
13. Royal College of Obstetricians and Gynaecologists, Post Hysterectomy vaginal Vault Prolapse (Green-top Guideline No.46) Addendum. Published February 2016. <https://www.rcog.org.uk/en/guidelines-research-services/guidelines/gtg46/>

Appendix A: Additional papers on sacrocolpopexy using mesh for vaginal vault repair

The following table outlines the studies that are considered potentially relevant to the IP overview but were not included in the main data extraction table (table 2). It is by no means an exhaustive list of potentially relevant studies. Given the large amount of evidence available on this procedure, non-randomised studies with less than 100 patients and follow-up less than 2 years were excluded from the analysis.

Article	Number of patients/follow-up	Direction of conclusions	Reasons for non-inclusion in table 2
Barber MD, Maher C (2013) Apical prolapse. International Urogynecology Journal 24:1815-1833.	Systematic review	Sacral colpopexy is an effective procedure for vault prolapse and further data are required on the route of performance and efficacy of this surgery for uterine prolapse. Polypropylene mesh is the preferred graft at ASC. Vaginal procedures for vault prolapse are well described and are suitable alternatives for those not suitable for sacral colpopexy.	Systematic review no meta-analysis. Non new safety events.
Bradley CS, Kenton KS, Richter HE et al. (2008) Obesity and outcomes after sacrocolpopexy. American journal of obstetrics and gynecology 199:1-8.	Prospective case series n=322 (74 obese, 122 overweight, 125 healthy weight) – in the original study FU=24 months	When compared to healthy weight women, obese women were younger, more likely to have stage 2 prolapse and had longer operative times. There was no difference in objective and subjective cure rates.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Chan SS, Pang SM, Cheung TH et al. (2011) Laparoscopic sacrocolpopexy for the treatment of vaginal vault prolapse: with or without robotic assistance. Hong Kong Medical Journal 17:54-60.	Retrospective case series n=36 (20 LSC versus 16 RASC) FU=29 months	Objective and subjective cure rates were similar in both groups. RASC was associated with longer hospital stay. There were no mesh erosions or exposure during follow-up.	Studies with more patients or longer follow-up are already included. Studies with more patients or longer follow-up are already included. No new safety events reported.
Claerhout F, De Ridder D, Roovers JP et al. (2008) Medium-term anatomic and functional results of laparoscopic	Prospective case series n=132 LSC	LSC demonstrated good objective and subjective cure rates. The posterior compartment was more	Studies with more patients or longer follow-up are already included.

sacrocolpopexy beyond the learning curve. European Urology 55:1459-67.	FU=13 months	vulnerable to prolapse at follow-up.	No new safety events reported.
Culligan PJ, Gurshumov E, Lewis C et al. (2014) Subjective and objective results 1 year after robotic sacrocolpopexy using a lightweight Y-mesh. International Urogynecology Journal 25:731-5.	Prospective case series n=150 RASC with lightweight mesh FU=12 months	Good subjective and objective cure rates. No mesh erosions or exposures at follow-up.	Studies with more patients or longer follow-up are already included. Included in the paper by Serati 2014.
Culligan PJ, Salamon C, Priestley JL, et al. (2013) Porcine dermis compared with polypropylene mesh for laparoscopic sacrocolpopexy: a randomized controlled trial. Obstetrics & Gynecology 121:143-151.	RCT n=115 LSC (57 Porcine graft versus 58 polypropylene mesh) FU=12 months	Similar objective and subjective cure rates in both groups. No major operative complications.	Study included in the paper by Maher 2013.
Cundiff GW, Varner E, Visco AG et al. (2008) Risk factors for mesh/suture erosion following sacral colpopexy. American journal of obstetrics and gynecology 688:1-5.	Prospective case series n=322 (157 SCP with Burch colposuspension versus 165 SCP only) – In the original RCT FU=2 years	Polytrafluoroethylene mesh was associated with higher rates of erosion and shouldn't be used for SCP. Concurrent hysterectomy and smoking are modifiable risk for mesh/suture erosion.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Deffieux X, Letouzey V, Savary D et al. (2012) Prevention of complications related to the use of prosthetic meshes in prolapse surgery: Guidelines for clinical practice. European Journal of Obstetrics Gynecology and Reproductive Biology 165:170-180.	Systematic review	The laparoscopic approach is recommended for sacral colpopexy (Expert opinion). It is recommended not to place and suture meshes by the vaginal route when a sacral colpopexy is performed (Grade B). It is recommended not to use silicone-coated polyester, porcine dermis, fascia lata, and polytetrafluoroethylene meshes (Grade B). It is recommended to use polyester (without silicone coating) or polypropylene meshes (Grade C). Suture of the meshes to the promontory can be performed using thread/needle or tacks (Grade C). Peritonization is recommended to cover the meshes (Grade C). If hysterectomy is required, it is recommended to perform a subtotal hysterectomy (Expert	Systematic review with no meta-analysis. No new safety events.

		opinion). Implementation of this guideline should decrease the prevalence of complications related to surgical procedures involving the use of prosthetic meshes.	
Deprest J, De Ridder D , Roovers JP et al. (2009) Medium-term anatomic and functional results of laparoscopic sacrocolpopexy beyond the learning curve. European Urology 55:1459-67.	Prospective case series n=150 (21 porcine grafts of small intestine submucosa, 29 dermal collagen versus 100 polypropylene mesh) FU=33 months	Overall anatomic failure was comparable. SCP using xenograft was associated with more apical failures and reoperations for prolapse than polypropylene.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Diwadkar GB, Barber MD, Feiner B (2009) Complication and reoperation rates after apical vaginal prolapse surgical repair: a systematic review. Obstetrics & Gynecology 113:367-373.	Systematic review.	The rate of complications requiring reoperation and the total reoperation rate was highest for vaginal mesh kits despite a lower reoperation rate for prolapse recurrence and shorter overall follow-up.	Systematic review with no meta-analysis. No new safety complications.
Filmar GA, Fisher HW, Aranda E et al. (2014) Laparoscopic uterosacral ligament suspension and sacral colpopexy: results and complications. International Urogynecology Journal 25:1645-1653.	Retrospective case series n=290 (102/290 stage 2 prolapse of which 73 LSC versus laparoscopic uterosacral ligament suspension) FU=112-114 days	There was no statistically significant difference in the rates of mesh erosion between concomitant total laparoscopic hysterectomy and prior hysterectomy. SCP resulted in statistically significant better anterior compartment support that uterosacral ligament suspension.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Ganatra AM, Rozet F, Sanchez-Salas R et al. (2009) The current status of laparoscopic sacrocolpopexy: a review. European Urology 55:1089-1103.	Systematic review	LSC upholds the outcomes of the gold standard ASC with minimal morbidity. Longer prospective and randomized trials are needed to confirm these results.	Systematic review with no meta-analysis. No new safety events reported.
Geller EJ, Parnell BA, Dunivan GC et al. (2012) Robotic vs abdominal sacrocolpopexy: 44-month pelvic floor outcomes. Urology 79:532-6.	Retrospective case series n=51 (23 RASC versus 28 SCP) FU=44 months	Objective and subjective success was similar in both groups. Mesh erosion rate was similar in both groups.	Studies with more patients or longer follow-up are already included. No new safety events reported.

Geller EJ, Siddiqui NY, Wu JM et al. (2008) Short-term outcomes of robotic sacrocolpopexy compared with abdominal sacrocolpopexy. <i>Obstetrics & Gynecology</i> 112:1201-1206.	Retrospective case series n=178 (73 RASC versus 105 SCP) FU=6 weeks	RASC demonstrated similar short-term vaginal vault support compared with SCP, with longer operative time, less blood loss and shorter length of stay.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Galczynski K, Nowakowski L, Romanek-Piva K et al. (2014) Laparoscopic mesh procedures for the treatment of pelvic organ prolapse--review of the literature. <i>Ginekologia Polska</i> 85:950-954.	Literature review. .	Laparoscopic sacrocolpopexy hysteropexy and lateral suspension are interesting and effective options for the treatment of pelvic organ prolapse, providing a number of important advantages characteristic for endoscopic techniques.	Literature review no meta-analysis.
Ginath S, Garely AD, Condrea A et al (2013) Mesh erosion following abdominal sacral colpopexy in the absence and presence of the cervical stump. <i>International Urogynecology Journal and Pelvic Floor Dysfunction</i> 24:113-118.	Retrospective case series n=277 (195 SCP with concomitant supracervical hysterectomy versus 82 SCP with previous total hysterectomy) FU=7-8 months	Similar objective success rates. Operative times were similar in both groups. The total hysterectomy group had higher rate of mesh erosion but this was not statistically significant follow-up.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Gupta P, Payne J , Killinger K A et al (2016) Analysis of changes in sexual function in women undergoing pelvic organ prolapse repair with abdominal or vaginal approaches. <i>International Urogynecology Journal and Pelvic Floor Dysfunction</i>	Prospective case series n= 204 FU= 12 months	74 out of 204 (36 %) had ASC and 130 out of 204 (64 %) had TVR. Seventy-two out of seventy-four ASCs were performed robotically and 2 were open. There was no difference in sexual activity or dyspareunia between the groups follow-up. PISQ and PFDI scores improved significantly in both the ASC and TVR groups over time compared with the baseline ($p < 0.0001$). Most women in the ASC (77.5 %) and TVR (64.8 %) groups were satisfied with the results of prolapse surgery at 12 months. Sexual function and pelvic floor symptoms improved in a similar manner in patients after abdominal and transvaginal POP surgery.	No new safety data, relevant outcomes. Non-randomised. Studies with larger follow-up already included.
Hill A J, Barber MD (2015) Apical prolapse repair: weighing the risks and benefits. <i>Current Opinion in Obstetrics & Gynecology</i> 27:373-379.	Systematic review.	Surgical restoration of the vaginal apex can be accomplished via a variety of approaches and techniques. When deciding on the proper surgical intervention, the surgeon must carefully	Systematic review with no meta-analysis. No new safety events.

		calculate the risks and benefits of each procedure while incorporating the patient's individual medical and surgical risk factors. Lastly, a discussion regarding the patient's overall goals of care is paramount to the decision-making process.	
Hudson CO, Northington GM, and Karp DR et al. (2012) Outcomes of robotic sacrocolpopexy: A systematic review and meta-analysis. Female pelvic medicine & reconstructive surgery 20: 252-260.	Systematic review and meta-analysis	RAASC is an effective surgical treatment for apical prolapse with high anatomic cure rate and low rate of complications.	Most of the reported studies overlap with papers analysed by Maher and Serati, both included in table 2. No new safety events reported.
Ivovic J, Kljakic D, and Raicevic (2012) Abdominal colposacropepy with permanent polypropylen mesh. Internet Journal of Gynecology and Obstetrics16:no pagination.	Prospective case series n=15 FU=9 months to 10 years	Satisfactory objective and subjective cure rates. No recurrence of prolapse, de novo urinary stress incontinence or dyspareunia.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Jeon MJ, Moon YJ, Jung HJ et al. (2009) A long-term treatment outcome of abdominal sacrocolpopexy. Yonsei Medical Journal 50:807-813.	Retrospective case series n=57 SCP FU=66 months	SCP had a reasonable objective cure rate for prolapse; nonetheless nearly half of the patients developed recurrent stress urinary incontinence within 1-3 months post-operatively. Bowel and sexual function did not significantly change after surgery. Major complications requiring reoperation or intensive care developed in one fifth of the patients. Two meshes were used: polytetrafluoroethylene and polypropylene. The author did not state how many patients had which type of mesh. Most severe complications happened in patients who had concomitant hysterectomy.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Jia X, Glazener C, Mowatt G (2010) Systematic review of the efficacy and safety of using mesh in surgery for uterine or vaginal vault prolapse. International Urogynecol Journal 21:1413-1431.	Systematic review	Sacrocolpopexy was associated with a low risk of recurrence but with a relatively high risk of mesh erosion. Ranges of estimates for outcomes for other mesh techniques were wide.	Systematic review no meta-analysis. No new safety events.
Khan A, Alperin M, Wu N et al. (2012) Comparative outcomes of open versus laparoscopic	Retrospective case series	LSC was associated with significantly higher rate of reoperation for anterior vaginal wall prolapse. More	Studies with more patients or longer follow-up are already included.

sacrocolpopexy among Medicare beneficiaries. International Urogynecology Journal 24:1883-1891.	n=970 (794 SCP versus 176 LSC) FU=12 months	medical complications (cardio-pulmonary) occurred post-operatively in the SCP group. When hysterectomy was done concomitantly, mesh related complications were significantly more frequent in the LSC group.	No new safety events reported.
Lamb SV, Massengill J, Sheridan MJ et al. (2015) Safety of combined abdominal sacral colpopexy and sigmoid resection with suture rectopexy: a retrospective cohort study. Female Pelvic Medicine & Reconstructive Surgery 21:18-24.	Retrospective case series n=194 (133 SCP, 34 SCP and sigmoid resection, 27 sigmoidectomy and rectopexy group) FU=12 months	The colorectal only group had a higher rate of ileus post-operatively. There were otherwise no differences in the rate of post-operative complications between groups. SCP using mesh at the time of sigmoid resection and anastomosis doesn't seem to increase the rate of post-operative complications.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Lee K, Motttrie A, Payne CK et al. (2014) A review of the current status of laparoscopic and robot-assisted sacrocolpopexy for pelvic organ prolapse. 65: 1128-1137.	Systematic review	LSC and RASC provide excellent short to medium term reconstructive outcomes for patients with POP. RASC is more expensive than LSC. Further studies are required to better understand the clinical performance of RASC versus LSC and confirm long-term efficacy.	Systematic review with no meta-analysis.
Leruth J, Fillet M, Waltregny D (2013) Incidence and risk factors of postoperative stress urinary incontinence following laparoscopic sacrocolpopexy in patients with negative preoperative prolapse reduction stress testing. International Urogynecology Journal and Pelvic Floor Dysfunction 24:485-491.	Retrospective case series n=55 LSC without concomitant SUI surgery after a negative preoperative prolapse reduction stress test FU=25 months	No patient developed recurrent prolapse or mesh erosion at follow-up. More than half of the patients reported symptoms of SUI postoperatively. Univariate analysis revealed that advanced cystocele (stage 3-4) and an history of patient-reported SUI before surgery were associated with higher risk of post-operative SUI after LSC. After 1 year follow-up, approximately one sixth of the patients underwent sling surgery.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Liang S, Zhu L, Song X et al. (2016) Long-term outcomes of modified laparoscopic sacrocolpopexy for advanced pelvic organ prolapse: a 3-year prospective study. Menopause 23:765-70.	Prospective case series n=30 modified LSC FU=3 years	Objective cure rates at 3 years follow-up were significant compared to baseline assessment. Sexual function was significantly improved. There was one case of mesh exposure and 2 cases of de novo dyspareunia.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Loffeld CJ, Thijs S, Mol BW et al. (2009) Laparoscopic sacrocolpopexy: a comparison of Prolene and Tutoplast mesh. Acta	Retrospective case series	There were no significant differences in operating time, blood loss or hospital stay between the groups. The risk of re-intervention because of	Studies with more patients or longer follow-up are already included.

Obstetricia et Gynecologica Scandinavica 88:826-830.	n=39 LSC (19 Tutoplast® versus 20 Prolene®) FU=45 months	prolapse was higher in the Tutoplast® group. Subjective cure rate was higher in the Prolene® group.	No new safety events reported.
Maheer CF, Feiner B, DeCuyper EM et al. (2011) Laparoscopic sacral colpopexy versus total vaginal mesh for vaginal vault prolapse: a randomized trial. American Journal of Obstetrics & Gynecology 204: 1-7.	RCT N 108 (53 LSC versus 55 TVM) FU=24 months	At two years the LSC had a higher satisfaction rate and objective success rate than the total vaginal mesh with lower perioperative morbidity and reoperation rate.	Study included in the paper by Maheer 2013 in table2.
Martin LA, Calixte R, Finamore PS (2015) Reoperation After Robotic and Vaginal Mesh Reconstructive Surgery: A Retrospective Cohort Study. Female Pelvic Medicine & Reconstructive Surgery 21:315-318.	Retrospective case series n=145 (181 RASC versus 64 transvaginal mesh repair) FU=RASC 3 months, TVM 12 months	TVM repair had shorter operation time. There was no significant difference in re-operation rate between the groups.	Studies with more patients or longer follow-up are already included. No new safety events reported.
McDermott CD, Park J, Terry CL et al. (2012) Surgical outcomes of abdominal versus laparoscopic sacral colpopexy related to body mass index. Journal of Obstetrics & Gynaecology Canada: JOGC 34:47-56.	Retrospective case series n=240 (90 SCP versus 150 LSC) FU=6-12 months	In normal weight patients, post-operative apical measurements were significantly worse in SCP patients. In overweight patients, the SCP group had significantly worse posterior measurements and fewer mesh erosions but more recurrent prolapse symptoms. In obese patients, the SCP group had significantly better anterior measurements. There was no significant difference between groups in regards to stage of prolapse, surgical satisfaction or surgical success or failure.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Mueller MG, Jacobs K M, Mueller ER et al. (2016) Outcomes in 450 Women After Minimally Invasive Abdominal Sacrocolpopexy for Pelvic Organ Prolapse. Female Pelvic Med Reconstr Surg 22:267-271.	Retrospective case series n=450 (232 SCP versus 226 RASC) FU=13 weeks	There were no significant differences between objective and subjective cure rates or bowel complications between the groups.	Studies with more patients or longer follow-up are already included. No new safety events reported.
North CE, Ali-Ross NS, Smith AR et al. (2009) A prospective study of laparoscopic sacrocolpopexy for the management of pelvic organ prolapse. BJOG: An	Retrospective case series n=22 LSC FU=27 months	Good rates of objective and subjective cure. Bowel symptoms were uncommon. Women have maintained sexual function with no dyspareunia.	Studies with more patients or longer follow-up are already included. No new safety events reported.

International Journal of Obstetrics & Gynaecology 116:1251-7.			
Porena M, Costantini E, Fioretti F et al. (2009) The management of pelvic organ prolapse: a review. Minerva Urologica e Nefrologica 61:363-371.	Systematic review.	SCP seems to obtain better anatomic outcomes than sacrospinous fixation but has a longer operation time and patient morbidity.	Systematic review with no meta-analysis. No new safety events.
Quiroz LH, Gutman RE, Shippey S et al (2008) Abdominal sacrocolpopexy: anatomic outcomes and complications with Pevicol, autologous and synthetic graft materials. American Journal of Obstetrics & Gynecology 198:1-5.	Retrospective case series n=259 LSC (102 Pevicol®, 134 synthetic mesh, 23 autologous fascia) FU=12 months	No statistically significant apical failure differences between groups. All reoperations occurred in the Pevicol® group. The Pevicol® groups had higher rates of mesh related complications but the difference wasn't statistically significant.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Rondini C, Braun H, Alvarez J et al. (2014) High uterosacral vault suspension vs Sacrocolpopexy for treating apical defects: a randomized controlled trial with twelve months follow-up. International Urogynecology Journal 26:1131-8.	RCT n=110 (54 SCP versus 56 high uterosacral vault suspension) FU=12 months	SCP has statistically better anatomical results when compared with HUVS for correcting apical defects at 12 months.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Salamon CG, Lewis C, Priestley J et al. (2013) Prospective study of an ultra-lightweight polypropylene Y mesh for robotic sacrocolpopexy. International Urogynecology Journal 24:1371-1375.	Prospective case series. n=120 RASC FU=12 months	Objective cure rates were satisfactory and subjective cure rates significant at follow-up. There were no mesh erosions or mesh related complication.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Sergent F, Resch B, Loisel C et al. (2011) Mid-term outcome of laparoscopic sacrocolpopexy with anterior and posterior polyester mesh for treatment of genito-urinary prolapse. European Journal of Obstetrics, Gynecology, and & Reproductive Biology 156:217-22.	Prospective case series. n=116 LSC FU=34 months	Anatomical success rates on the apical, anterior or posterior compartments were respectively, 97%, 89% and 98%. On the functional level all the scores of quality of life and sexuality were improved.	Some patients had concomitant hysterectomy alongside LSC and it wasn't possible to obtain sub-group analysis for the outcomes of interest of this synthesis.
Shimko MS, Umbreit EC, Chow GK et al. (2011) Long-term outcomes of robotic-assisted laparoscopic sacrocolpopexy with a minimum of three years follow-up. Journal of robotic surgery 5: 175-180.	n= 40 FU= 62 months	RASC was associated with a short hospital stay, low complication rates, and high patient satisfaction with a minimum of 3 years' follow-up.	Larger studies with greater follow-up already included. Relevant outcome data and long-term follow-up.

Sierra JM , Oshiro EO, Perez CF et al. (2011) Long-term outcomes after robotic sacrocolpopexy in pelvic organ prolapse: prospective analysis. Urologia Internationalis 86:414-418.	Prospective case series n=31 RASC FU=25 months	There was 1 conversion to SCP. There were 2 major complications (1 acute myocardial infarction and 1 reoperation for excess tension with syncope), two minor complications (1 wound infection and 1 ileus) and no recurrences at follow-up. Success of RASC might improve with experience. More evidence is required regarding safety of RASC.	Already included in the paper by Serati (2014).
Stanford EJ, Cassidenti A, Moen MD (2012) Traditional native tissue versus mesh-augmented pelvic organ prolapse repairs: providing an accurate interpretation of current literature. International Urogynecology Journal 23:19-28.	Systematic review.	There may be a higher rate of complications noted for mesh implantation. POP surgery is complex, and both NT and MA techniques require skills to perform proper compartmental reconstruction. An understanding of the published literature and knowledge of individual surgeon factors are important in deciding which surgical approach to use and how to best counsel patients during informed consent.	Systematic review no meta-analysis.
Tan-Kim J, Menefee SA, Lippmann Q et al. (2014) A pilot study comparing anatomic failure after sacrocolpopexy with absorbable or permanent sutures for vaginal mesh attachment. Permanente Journal 18:40-44.	Retrospective case series n=193 SCP (45 delayed absorbable sutures versus 148 permanent sutures) FU=43 weeks	Objective failure rates differences were not statistically significant for all compartments.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Thomas AZ, Giri SK, Cox AM et al. (2009) Long-term quality-of-life outcome after mesh sacrocolpopexy for vaginal vault prolapse. BJU International 104: 1676-1679.	Prospective case series n=21 SCP FU=52.2 months	Total PDFI scores were significantly better after SCP. All patients reported a significant improvement of symptoms in the POPDI category. CRADI subscale score showed no significant change after SCP. There was an improvement of urinary symptoms in the UDI subscale but this wasn't statistically significant. Analysis of score differences over time after SCP showed an insignificant decreasing slope. Suggestion of long-term stability of symptoms, improved sexual function and patient satisfaction.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Turner L, Lavelle E, Lowder JL et al. (2016) The Impact	Case series	In women undergoing minimally invasive	Large percentage of patients having

of Obesity on Intraoperative Complications and Prolapse Recurrence after Minimally Invasive Sacrocolpopexy. Female Pelvic Medicine and Reconstructive Surgery 22(5), 317-323.	n=556 6 weeks	sacrocolpopexy, obesity is associated with increased blood loss, longer operative times, and more intraoperative complications, specifically conversions to laparotomy. Even after correcting for blood loss, surgeon experience, and concomitant hysterectomy, obese women were 3 times as likely to have an intraoperative complication. Our data did not show that obesity was associated with increased risk of prolapse recurrence; however, postoperative follow-up was limited.	concomitant hysterectomy. Does report on a rare safety event: corneal abrasion.
Unger CA, Paraiso MF, Jelovsek JE et al. (2014) Perioperative adverse events after minimally invasive abdominal sacrocolpopexy. American Journal of Obstetrics & Gynecology 211:1-8.	Retrospective case series n= 406 (249 LSC versus RASC 121) FU=7 months	RSC was significantly associated with higher intra-operative bladder injury rate, higher estimated blood loss and reoperation rate for pelvic organ prolapse compared with LSC. Concomitant rectopexy was significantly associated with a higher risk of transfusion, pelvic/abdominal abscess formation and osteomyelitis. Mesh erosion rate difference didn't reach statistical significance.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Wong V, Guzman-Rojas R, Shek KL et al. (2016) Laparoscopic sacrocolpopexy: how low does the mesh go? Ultrasound Obstet: DOI: 10.1002/uog.15882Gynecol	Retrospective case series n= 97 LSC FU=3 years	At follow-up there were no recurrences of apical prolapse but cystocele recurrence was common despite emphasis on anterior mesh extension. The author suggests that the lower the mesh reaches towards the bladder neck, the less likely will anterior compartment occur.	Studies with more patients or longer follow-up are already included. No new safety events reported.
Yurteri-Kaplan LA, Gutman RE (2012) The use of biological materials in urogynecologic reconstruction: a systematic review. Plastic & Reconstructive Surgery 130:242S-253S.	Systematic review	For prolapse surgery, the addition of a biological graft adds no benefit compared with native tissue repairs for rectocele repair. Conflicting data exist regarding cystocele repair. Synthetic mesh repairs provide superior anatomical support for sacral colpopexy and cystocele repair compared with biologic grafts. However, biological and synthetic mesh slings have equivalent success rates for the treatment of stress urinary incontinence. Contrary to prior assumptions that biologic	Systematic review with no meta-analysis. No new safety report.

		grafts add tissue strength without graft-related complications, there appears to be no benefit to the use of biological materials for prolapse and incontinence surgery.	
--	--	--	--

Appendix B: Related NICE guidance for sacrocolpopexy using mesh to repair vaginal vault prolapse

Guidance	Recommendations
Interventional procedures	<p>Sacrocolpopexy with hysterectomy using mesh to repair uterine prolapse. NICE interventional procedure guidance IPG577 (2017).</p> <p>1.1 Current evidence on the safety and efficacy of sacrocolpopexy with hysterectomy using mesh to repair uterine prolapse is inadequate in quantity and quality. Therefore this procedure should only be used with special arrangements for clinical governance, consent and audit or research.</p> <p>1.2 Clinicians wishing to do sacrocolpopexy with hysterectomy using mesh to repair uterine prolapse should:</p> <ul style="list-style-type: none"> • Inform the clinical governance leads in their trusts. • During the consent process, ensure that patients understand the uncertainty about the procedure's safety, including mesh erosion (for example, into the vagina) and the risk of recurrence, and provide them with clear written information. In addition, the use of NICE's information for the public is recommended. <p>1.3 Patient selection and treatment should only be done by specialists with experience in managing pelvic organ prolapse and urinary incontinence in women. All clinicians doing this procedure should have specific up-to-date training in the procedure.</p> <p>1.4 Clinicians should enter details about all patients having sacrocolpopexy with hysterectomy using mesh to repair uterine prolapse onto an appropriate registry (for example, the British Society of Urogynaecology database). All adverse events involving the medical device used in this procedure should be reported to the Medicines and Healthcare products Regulatory Agency.</p> <p>1.5 NICE may update the guidance on publication of further evidence.</p> <p>Single-incision short sling mesh insertion for stress urinary incontinence in women. NICE interventional procedure guidance IPG566 (2016).</p> <p>1.1 The evidence on the safety of single-incision short sling mesh insertion for stress urinary incontinence in women shows infrequent but serious complications. These include lasting pain, discomfort and failure of the procedure. The mesh implant is intended to be</p>

	<p>permanent but, if removal is needed because of complications, the anchoring system can make the device very difficult or impossible to remove. The evidence on efficacy in the long term is inadequate in quality and quantity. Therefore, this procedure should not be used unless there are special arrangements in place for clinical governance, consent, and audit or research.</p> <p>1.2 Clinicians wishing to do single-incision short sling mesh insertion for stress urinary incontinence in women should:</p> <ul style="list-style-type: none"> • Inform the clinical governance leads in their NHS trusts. • Ensure that patients understand the uncertainty about the procedure's safety and efficacy, including that there is the potential for the procedure to fail and for serious long-term complications from the device, and that the mesh implant is intended to be permanent so removal, if needed, may be difficult or impossible. Provide patients with clear written information. In addition, the use of NICE's information for the public is recommended. • Audit and review clinical outcomes of all patients having single-incision short sling mesh insertion for stress urinary incontinence in women (see section 7.1). <p>1.3 Patient selection should be done by a multidisciplinary team with experience in the assessment and management of women with stress urinary incontinence.</p> <p>1.4 This procedure should only be done by clinicians with specific training in transobturator surgical techniques. Removal of a short sling mesh should only be done by people with expertise in this specialised surgery.</p> <p>1.5 NICE encourages further research into single-incision short sling mesh insertion for stress urinary incontinence in women and may update the guidance on publication of further evidence. Studies should include details of patient selection, and should measure long-term outcomes including effects on quality of life and other patient-reported outcomes.</p> <p>Insertion of mesh uterine suspension sling (including sacrohysteropexy) for uterine prolapse repair. NICE Interventional procedure guidance IPG282 (2009).</p> <p>1.1 Current evidence on the safety and efficacy of insertion of mesh uterine suspension sling (including sacrohysteropexy) for uterine prolapse repair is inadequate in quantity. Therefore this procedure should only be used with special arrangements for clinical governance, consent and audit or research.</p>
--	---

	<p>1.2 Clinicians wishing to undertake insertion of mesh uterine suspension sling (including sacrohysteropexy) for uterine prolapse repair should take the following actions.</p> <ul style="list-style-type: none"> • Inform the clinical governance leads in their Trusts. • Ensure that patients understand the uncertainty about the procedure's safety, including mesh erosion (for example, into the vagina) and the risk of recurrence, and provide them with clear written information. In addition, use of NICE's information for patients ('Understanding NICE guidance') is recommended. <p>1.3 The procedure should only be carried out by surgeons specialising in the management of pelvic organ prolapse and female urinary incontinence.</p> <p>1.4 The British Society for Urogynaecology runs a database on urogynaecological procedures, and clinicians should enter details about all patients undergoing this procedure onto this database.</p> <p>1.5 NICE encourages further research into mesh uterine suspension sling (including sacrohysteropexy) for uterine prolapse repair and may review the procedure on publication of further evidence on different types of mesh. Future research should include short- and long-term efficacy, safety outcomes (such as mesh erosion in the long term), patient-reported quality-of-life outcomes using validated scales and subsequent successful pregnancy.</p> <p>Infracoccygeal sacropexy using mesh for vaginal vault prolapse repair. NICE interventional procedure guidance IPG281 (2009).</p> <p>1.1 Current evidence on the efficacy and safety of infracoccygeal sacropexy using mesh for vaginal vault prolapse repair is inadequate in quantity and quality. Therefore this procedure should only be used with special arrangements for clinical governance, consent and audit or research.</p> <p>1.2 Clinicians wishing to undertake infracoccygeal sacropexy using mesh for vaginal vault prolapse repair should take the following actions:</p> <ul style="list-style-type: none"> • Inform the clinical governance leads in their Trusts. • Ensure that patients understand the uncertainty about the procedure's safety, including mesh erosion (for example, into the vagina) and the risk of recurrence, and provide them with clear written information. In addition, use of NICE's information for patients ('Understanding NICE guidance') is recommended. <p>1.3 The procedure should only be carried out by surgeons specialising in the management of pelvic organ prolapse and female urinary incontinence.</p> <p>1.4 The British Society for Urogynaecology runs a database on urogynaecological procedures, and clinicians should enter details about all patients undergoing this procedure onto this database.</p> <p>1.5 NICE encourages further research into infracoccygeal sacropexy using mesh for vaginal vault prolapse repair, and may review the</p>
--	--

	<p>procedure on publication of further evidence on different types of mesh. Clinicians are encouraged to collect long-term data on clinical outcomes and patient-reported quality-of-life outcomes using validated scales.</p> <p>Infracoccygeal sacropexy using mesh for uterine prolapse repair. NICE Interventional procedure guidance IPG280 (2009).</p> <p>1.1 Current evidence on the efficacy and safety of infracoccygeal sacropexy using mesh for uterine prolapse repair is inadequate in quantity and quality. Therefore this procedure should only be used with special arrangements for clinical governance, consent and audit or research.</p> <p>1.2 Clinicians wishing to undertake infracoccygeal sacropexy using mesh for uterine prolapse repair should take the following actions:</p> <ul style="list-style-type: none"> • Inform the clinical governance leads in their Trusts. • Ensure that patients understand the uncertainty about the procedure's safety, including mesh erosion (for example, into the vagina) and the risk of recurrence, and provide them with clear written information. In addition, use of NICE's information for patients ('Understanding NICE guidance') is recommended. <p>1.3 The procedure should only be carried out by surgeons specialising in the management of pelvic organ prolapse and female urinary incontinence.</p> <p>1.4 The British Society for Urogynaecology runs a database on urogynaecological procedures, and clinicians should enter details about all patients undergoing this procedure onto this database.</p> <p>1.5 NICE encourages further research into infracoccygeal sacropexy using mesh for uterine prolapse repair, and may review the procedure on publication of further evidence on different types of mesh. Clinicians are encouraged to collect long-term data on clinical outcomes and patient-reported quality-of-life outcomes using validated scales.</p> <p>Surgical repair of vaginal wall prolapse using mesh. NICE Interventional procedure guidance IPG267 (2008).</p> <p>1.1 The evidence suggests that surgical repair of vaginal wall prolapse using mesh may be more efficacious than traditional surgical repair of vaginal wall prolapse without mesh. Both efficacy and safety vary with different types of mesh, and the data on efficacy in the long term are limited in quantity. There is a risk of complications that can cause significant morbidity. Therefore, this procedure should only be used with special arrangements for clinical governance, consent and audit or research.</p> <p>1.2 Clinicians wishing to undertake surgical repair of vaginal wall prolapse using mesh should take the following actions.</p> <ul style="list-style-type: none"> • Inform the clinical governance leads in their Trusts.
--	--

	<ul style="list-style-type: none"> • Ensure that patients understand that there is uncertainty about the long-term results and there is a risk of complications, including sexual dysfunction and erosion into the vagina, which would require additional procedures. They should provide them with clear written information. In addition, the use of the Institute's information for patients ('Understanding NICE guidance') is recommended. • Audit and review clinical outcomes of all patients having surgical repair of vaginal wall prolapse using mesh (see section 3.1). <p>1.3 This is a technically challenging procedure that should only be carried out by gynaecologists with special expertise in the surgical management of pelvic organ prolapse. Specific training is required when trocar introducer systems are used for the insertion of mesh.</p> <p>1.4 Further publication of safety and efficacy outcomes will be useful. Research should aim to address the performance of different methods of repair and different types of mesh. It should also include evidence about long-term outcomes and patient-reported outcomes, such as quality of life and sexual function. The Institute may review the procedure upon publication of further evidence.</p>
NICE guidelines	<p>Urinary incontinence in women (2013) NICE guideline CG171 (2013). 1.10 Surgical approaches for SUI</p> <p>1.10.1 When offering a surgical procedure discuss with the woman the risks and benefits of the different treatment options for SUI using the information in information to facilitate discussion of risks and benefits of treatments for women with stress urinary incontinence. [new 2013]</p> <p>1.10.2 If conservative management for SUI has failed, offer:</p> <ul style="list-style-type: none"> • synthetic mid-urethral tape (see recommendations 1.10.3–8), or • open colposuspension (see also recommendation 1.10.9), or • autologous rectus fascial sling (see also recommendation 1.10.10). [new 2013] <p>Synthetic tapes</p> <p>1.10.3 When offering a synthetic mid-urethral tape procedure, surgeons should:</p> <ul style="list-style-type: none"> • use procedures and devices for which there is current high quality evidence of efficacy and safety^[10] • only use a device that they have been trained to use (see recommendations in section 1.11) • use a device manufactured from type 1 macroporous polypropylene tape • consider using a tape coloured for high visibility, for ease of insertion and revision. [new 2013]

	<p>1.10.4 If women are offered a procedure involving the transobturator approach, make them aware of the lack of long-term outcome data. [new 2013]</p> <p>1.10.5 Refer women to an alternative surgeon if their chosen procedure is not available from the consulting surgeon. [new 2013]</p> <p>1.10.6 Use 'top-down' retropubic tape approach only as part of a clinical trial. [new 2013]</p> <p>1.10.7 Refer to single-incision sub-urethral short tape insertion for stress urinary incontinence (NICE interventional procedure guidance 262) for guidance on single-incision procedures. [new 2013]</p> <p>1.10.8 Offer a follow-up appointment (including vaginal examination to exclude erosion) within 6 months to all women who have had continence surgery. [new 2013]</p> <p>Colposuspension</p> <p>1.10.9 Do not offer laparoscopic colposuspension as a routine procedure for the treatment of stress UI in women. Only an experienced laparoscopic surgeon working in an MDT with expertise in the assessment and treatment of UI should perform the procedure. [2006]</p> <p>Biological slings</p> <p>1.10.10 Do not offer anterior colporrhaphy, needle suspensions, paravaginal defect repair and the Marshall–Marchetti–Krantz procedure for the treatment of stress UI. [2006]</p> <p>Intramural bulking agents</p> <p>1.10.11 Consider intramural bulking agents (silicone, carbon-coated zirconium beads or hyaluronic acid/dextran copolymer) for the management of stress UI if conservative management has failed. Women should be made aware that:</p> <ul style="list-style-type: none"> • repeat injections may be needed to achieve efficacy • efficacy diminishes with time • efficacy is inferior to that of synthetic tapes or autologous rectus fascial slings. [2006, amended 2013] <p>1.10.12 Do not offer autologous fat and polytetrafluoroethylene used as intramural bulking agents for the treatment of stress UI. [2006]</p> <p>Artificial urinary sphincter</p> <p>1.10.13 In view of the associated morbidity, the use of an artificial urinary sphincter should be considered for the management of stress UI in women only if previous surgery has failed. Life-long follow-up is recommended. [2006]</p> <p>Considerations following unsuccessful invasive SUI procedures or recurrence of symptoms</p> <p>1.10.14 Women whose primary surgical procedure for SUI has failed (including women whose symptoms have returned) should be:</p> <ul style="list-style-type: none"> • referred to tertiary care for assessment (such as repeat urodynamic testing including additional tests such as imaging
--	--

	<p>and urethral function studies) and discussion of treatment options by the MDT, or</p> <ul style="list-style-type: none">• offered advice as described in recommendation 1.6.9 if the woman does not want continued invasive SUI procedures. <p>[new 2013]</p>
--	--

Appendix C: Literature search for sacrocolpopexy using mesh for vaginal vault prolapse repair

Databases	Date searched	Version/files
Cochrane Database of Systematic Reviews – CDSR (Cochrane Library)	28/11/2016	Issue 11 of 12, November 2016
Cochrane Central Database of Controlled Trials – CENTRAL (Cochrane Library)	28/11/2016	Issue 10 of 12, October 2016
HTA database (Cochrane Library)	28/11/2016	Issue 4 of 4, October 2016
MEDLINE (Ovid)	28/11/2016	1946 to November week 3 2016
MEDLINE In-Process (Ovid)	28/11/2016	November 22, 2016
EMBASE (Ovid)	28/11/2016	1974 to 2016 Week 47
PubMed	28/11/2016	N/A
JournalTOCS	28/11/2016	N/A

Trial sources searched on 09 06 2016

- Clinicaltrials.gov
- ISRCTN
- WHO International Clinical Trials Registry

Websites searched on 09 06 2016

- National Institute for Health and Care Excellence (NICE)
- NHS England
- Food and Drug Administration (FDA) - MAUDE database
- Australian Safety and Efficacy Register of New Interventional Procedures – Surgical (ASERNIP – S)
- Australia and New Zealand Horizon Scanning Network (ANZHSN)
- EuroScan
- General internet search

The following search strategy was used to identify papers in MEDLINE. A similar strategy was used to identify papers in other databases.

- 1 pelvic organ prolapse/
- 2 POP.ti,ab.
- 3 Uterine Prolapse/
- 4 vagina/

IP overview: Sacrocolpopexy using mesh to repair vaginal vault prolapse

- 5 fascia/
- 6 ((apical* or post-hysterect* or cuff* or fascia* or pelvic* or cervic* or transvagin* or vagin* or genital* or uter* or urogenit* or womb* or genito* or intravaginal*) adj2 (prolaps* or collaps* or drop*)).ti,ab.
- 7 rectocele/
- 8 cystocele/
- 9 (rectocele* or cystocele* or enterocele*).ti,ab.
- 10 or/1-9
- 11 surgical mesh/
- 12 mesh*.ti,ab.
- 13 (biologic* adj4 (graft* or plast* or sling* or tape* or suspens* or gauze*)).ti,ab.
- 14 *Polypropylenes/ or *Polyglactin 910/
- 15 ((Polypropylene* or Polyglactin* or Novasilk* or Restonelle* or prolene* or trelex* or avaulta* or pelvitex* or prolift* or polyform* or marlex* or gynemesh* or gore* or vicryl* or tutoplast* or faslata* or fortagen* or porcine dermis* or pelvicol* or pelvisoft* or upsylon* or Elevate PC or bovine pericardium) adj2 (mesh* or graft* or plast* or sling* or tape* or suspens* or gauze*)).ti,ab.
- 16 or/11-15
- 17 10 and 16
- 18 *gynecologic surgical procedures/
- 19 suburethral slings/
- 20 urogenital surgical procedures/ or urologic surgical procedures/
- 21 (Colporrhaph* or colpoperineorrhaph* or cystopex* or sacrohysteropex* or sacrocolpopex* or sacropex*).ti,ab.
- 22 or/18-21
- 23 17 and 22
- 24 (artisyn Y-shaped or inte-pro Y or uplift or prolife or perigee or apogee or elevate or capio or avaulta or i-stitch or restorelle or uphold LITE).ti,ab.
- 25 10 and 24

- 26 23 or 25
- 27 animals/ not humans/
- 28 26 not 27
- 29 limit 28 to ed=20070701-20160630