

ENDURALIFE powered CRT-D devices for treating heart failure

Medical technologies guidance

Published: 17 March 2017

Last updated: 21 July 2021

www.nice.org.uk/guidance/mtg33

Your responsibility

This guidance represents the view of NICE, arrived at after careful consideration of the evidence available. When exercising their judgement, healthcare professionals are expected to take this guidance fully into account, and specifically any special arrangements relating to the introduction of new interventional procedures. The guidance does not override the individual responsibility of healthcare professionals to make decisions appropriate to the circumstances of the individual patient, in consultation with the patient and/or guardian or carer.

All problems (adverse events) related to a medicine or medical device used for treatment or in a procedure should be reported to the Medicines and Healthcare products Regulatory Agency using the [Yellow Card Scheme](#).

Commissioners and/or providers have a responsibility to implement the guidance, in their local context, in light of their duties to have due regard to the need to eliminate unlawful discrimination, advance equality of opportunity, and foster good relations. Nothing in this guidance should be interpreted in a way that would be inconsistent with compliance with those duties. Providers should ensure that governance structures are in place to review, authorise and monitor the introduction of new devices and procedures.

Commissioners and providers have a responsibility to promote an environmentally sustainable health and care system and should [assess and reduce the environmental impact of implementing NICE recommendations](#) wherever possible.

Contents

1 Recommendations	4
2 The technology.....	5
Description of the technology	5
Current management.....	6
3 Clinical evidence	8
Summary of clinical evidence	8
4 NHS considerations	18
System impact.....	18
5 Cost considerations	19
Cost evidence.....	19
6 Conclusions	27
7 Committee members and NICE project team.....	28
Committee members	28
NICE project team	28
Update information	29

1 Recommendations

- 1.1 The case for adopting ENDURALIFE-powered cardiac resynchronisation therapy-defibrillator (CRT-D) devices for treating heart failure is supported by the published evidence. Extended battery life is of clinical and patient benefit and associated with fewer replacement procedures.
- 1.2 ENDURALIFE-powered CRT-Ds should be considered as an option in people offered CRT-D devices in line with [NICE technology appraisal guidance on implantable cardioverter defibrillators and cardiac resynchronisation therapy for arrhythmias and heart failure](#).
- 1.3 Cost modelling was based on published data using predecessor devices, and showed that the price and lifespan of the CRT-D have the greatest effect on overall treatment costs. Assuming an average selling price of £15,322, across different devices, using ENDURALIFE-powered CRT-Ds may save between £2,614 and £6,941 per patient using NHS tariff costs and between £2,313 and £6,140 per patient using NHS reference costs over 15 years through a reduction in the need for replacement procedures [2021]. This could save the NHS in England around £6 million in the first 5 years.

2 The technology

Description of the technology

- 2.1 The ENDURALIFE battery technology (Boston Scientific) is designed to extend the battery life of Boston Scientific cardiac resynchronisation therapy-defibrillator (CRT-D) devices. CRT-Ds are a treatment option for heart failure and life-threatening ventricular arrhythmias. ENDURALIFE battery technology uses a lithium manganese dioxide (Li/MnO₂) battery chemistry, which is claimed to be less susceptible to the variations in voltage and resistance associated with early battery depletion. CRT-Ds with ENDURALIFE battery technology are also claimed to use less current than other CRT-Ds.
- 2.2 ENDURALIFE battery technology was first incorporated into the COGNIS CRT-D device in February 2008. The ENDURALIFE brand was launched in 2015, but the technology has been in all Boston Scientific CRT-Ds since 2008; the CE marks for the RESONATE, MOMENTUM, CHARISMA, VIGILANT, AUTOGEN, DYNAGEN, INOGEN, ORIGEN, INCEPTA, ENERGEN, PUNCTUA and COGNIS CRT-Ds all include the ENDURALIFE battery technology. ENDURALIFE-powered CRT-Ds are CE-marked as class III medical devices.
- 2.3 According to the company's submission, the cost model uses an average industry-wide cost of £12,404 for a complete CRT-D system. This cost was derived from the average selling prices used in the economic modelling for [NICE technology appraisal guidance on implantable cardioverter defibrillators and cardiac resynchronisation therapy for arrhythmias and heart failure](#). The cost was inflated using the 2015 Bank of England inflation rate of 0.9%.
- 2.4 The claimed benefits of ENDURALIFE-powered CRT-Ds in the case for adoption presented by the company are:

- Longer battery life for devices could help improve patient experience by increasing the time between replacements (meaning fewer overall replacement surgeries).
- Fewer CRT-D replacements could be particularly beneficial for patients with heart failure who are already very unwell and may have difficulty lying down for a long time.
- Fewer replacements will also reduce the risk of complications, which is higher in replacement procedures than in primary implants. The increased risk of complications and infections can have a measurable impact on morbidity and mortality.
- Fewer early replacements will lead to savings for the healthcare system, such as a reduction in hospital admissions, bed days and procurement costs. Fewer replacement procedures also means a reduction in associated costs such as post-operative complications and infections. Preliminary estimates suggest it could represent £33 million over 6 years.
- Reduced replacement rates will allow more new patients to have implants within the same resource constraints, thus supporting the implementation of [NICE's technology appraisal guidance on implantable cardioverter defibrillators and cardiac resynchronisation therapy for arrhythmias and heart failure](#) and helping to meet the recommended levels of CRT-D implants in the UK.

Current management

- 2.5 [NICE has published a guideline on the diagnosis and management of chronic heart failure in adults. NICE technology appraisal guidance on implantable cardioverter defibrillators and cardiac resynchronisation therapy for arrhythmias and heart failure](#) recommends CRT-Ds as an adjunctive treatment option for heart failure in people on optimal medical therapy who have left ventricular dysfunction with a left ventricular ejection fraction of 35% or less.
- 2.6 Implantation of an ENDURALIFE-powered device uses standard CRT-D insertion techniques. Expert advisers have stated that people with a CRT-D are typically followed-up by a physiologist in a technical device clinic and either a routine cardiology or specialist heart failure clinic.

Patients with a CRT-D usually attend a technical device clinic every 3 months, unless remote telemonitoring is used. It is recommended that patients should have one face-to-face technical device review annually. Patients will also need to be seen by a cardiologist; these clinics are dictated by clinical need/patient stability but are usually 6-monthly. When possible, the aim is to coincide the technical and cardiology clinics once a year. At each attendance, the patient's clinical status is noted and the device interrogated. Tests include the pacing function, the defibrillation leads (including lead impedance), the time spent pacing and the incidence of arrhythmias. The rate of battery depletion, and therefore the anticipated lifespan of the device, is also noted.

- 2.7 Remote device monitoring systems, which may reduce the need for technical device attendances, are available for all CRT-Ds, including those with ENDURALIFE battery technology.

3 Clinical evidence

Summary of clinical evidence

- 3.1 The key clinical outcomes for ENDURALIFE-powered cardiac resynchronisation therapy-defibrillator (CRT-D) devices in the decision problem were:
- device survival
 - battery survival (or time to battery depletion)
 - CRT-D component failure
 - number of invasive procedures including CRT-D replacements
 - incidence of complications after replacement procedures for battery depletion or CRT-D component failure (as per definitions in the REPLACE registry)
 - inpatient admissions and bed days (related to interventions)
 - death
 - patient satisfaction
 - quality of life
 - device-related adverse events.
- 3.2 The company did 2 searches for published literature on studies of device lifespan, the incidence of complications associated with device replacement, and outcomes relating to patient quality of life or satisfaction associated with device replacement. Its submission included: 6 case series studies of CRT-D lifespan reported in 7 sources; 5 product performance reviews (see [section 3.9](#)); and 20 studies (17 observational studies and 3 systematic reviews) that highlight the complications associated with implantable cardioverter defibrillators or CRT-D replacement, as well as patient preference for device size or lifespan. The external assessment centre (EAC) excluded 14 of the

17 observational studies because data from these studies were also used in the submitted systematic reviews. It judged that 1 further study on complications (Kirkfeldt et al. 2014), identified by clinical experts, was relevant. In total, the EAC assessed 6 observational studies on ENDURALIFE-powered CRT-D battery life (Alam et al. 2016, Ellis et al. 2016, Landolina et al. 2015, Von Gunten et al. 2015, Lau et al. 2015 and Williams and Stevenson 2014), 5 product performance reviews (Boston Scientific, Biotronik, Medtronic, Sorin and St Jude Medical) and 6 studies on adverse events arising from cardiac device replacement (Lewis et al. 2016, Polyzos et al. 2015, Zeitler et al. 2015, Nichols et al. 2016, Lovelock et al. 2014 and Kirkfeldt et al. 2014).

Battery life

3.3 Alam et al. (2016) and Alam et al. (2014) are retrospective observational studies, both reporting on the same cohort, evaluating the time from device implantation to battery depletion. The most recent publication included 621 patients, of which 122 had ENDURALIFE-powered CRT-Ds, 51 had a non-ENDURALIFE-powered Boston Scientific device and 448 had a device from another company (Medtronic n=391, St Jude Medical n=57). The devices were implanted between January 2008 and December 2010, with a maximum possible follow-up of 8 years and mean follow-up was 3.4 years. Rates of CRT-D replacement because of battery depletion were 16% (Boston Scientific) compared with 51% to 53% for devices from other companies. When comparing time to battery depletion, Boston Scientific devices lasted longer than either Medtronic (hazard ratio [HR] 0.15, 95% confidence interval [CI] 0.10 to 0.22, $p < 0.001$) or St Jude Medical devices (HR 0.28, 95% CI 0.16 to 0.48, $p < 0.001$). The hazard ratios for battery depletion (adjusted for unbalanced electrical pacing parameters) were:

- Boston Scientific compared with Medtronic: 0.11 (95% CI 0.07 to 0.16, $p < 0.001$)
- Boston Scientific compared with St Jude Medical: 0.25 (95% CI 0.13 to 0.47, $p < 0.001$).

Of the 67 patients still alive 6 years after implantation, battery survival rates were 77% (Boston Scientific), 44% (St Jude Medical) and 10% (Medtronic).

- 3.4 Ellis et al. (2016) is a retrospective observational study designed to assess how the battery capacity of a CRT-D affects the time until the elective replacement indicator (ERI) is reached. A total of 1,302 CRT-Ds (Boston Scientific n=322 (97.0% ENDURALIFE-powered CRT-Ds), Medtronic n=794 and St Jude Medical n=186) were implanted between August 2008 and December 2010. Over a mean follow-up of 3 years, the proportions of devices reaching ERI were: 0.3% (Boston Scientific, battery capacity=2.0 Ah), 13.5% (Medtronic, 1.0 Ah) and 3.8% (St Jude Medical, 1.4 Ah). The odds ratio (OR) for reaching ERI with a Medtronic device (1.0 Ah) compared with a St Jude Medical (1.4 Ah) or Boston Scientific (2.0 Ah) device was 9.73 ($p < 0.0001$). Univariate predictors for ERI included 1.0 Ah device and an LV pacing output of over 3 V at 1 ms (OR: 3.74, $p < 0.001$). Mortality rates with each device were 28.0% (Boston Scientific), 16.7% (St Jude Medical) and 21.8% (Medtronic). No CRT-D failures were observed. High left ventricle lead impedance was protective of reaching ERI: OR (>1,000 versus 500 Ohms) 0.38, 95% CI 0.20 to 0.71, $p = 0.0025$.
- 3.5 Landolina et al. (2015) is a retrospective observational study examining the rate of replacement for battery depletion and to identify reasons for early depletion. A total of 1,726 CRT-Ds (Boston Scientific n=608 [291 (47.9%) ENDURALIFE-powered CRT-Ds], Biotronik n=49, Sorin n=99, St Jude Medical n=172 and Medtronic n=798) were implanted from January 2008 to March 2010. The CRT-Ds were commercially released between 2003 and 2010 and had different battery types; 708 were early-generation (released before 2007) and 1,018 were recent-generation families (since 2007). The median follow-up was 3.6 years. Among the recent-generation CRT-Ds (excluding those from Sorin and Biotronik, because there were fewer than 100 of these implants included in the study), rates of devices still working after 5 years were 88% (Boston Scientific), 75% (St Jude Medical) and 52% (Medtronic). Table 1 shows multivariate analysis factors associated with CRT-D replacement because of battery depletion.

Table 1 Factors associated with CRT-D replacement because of battery depletion

Factors associated with battery depletion	Hazard ratio	95% confidence interval	P value
Boston Scientific vs Medtronic	0.64	0.47–0.89	0.008
Recent-generation device	0.57	0.45–0.72	<0.001
Battery chemistry: Li/MnO ₂ vs Li/SVO	0.37	0.22–0.64	<0.001
Li/CF _x -SVO vs Li/SVO	0.28	0.16–0.50	<0.001
High left ventricle lead output (pulse amplitude >2.5 V, duration >0.5 ms)	1.96	1.57–2.46	<0.001
Unipolar left ventricular lead	1.58	1.25–2.01	<0.001

- 3.6 Von Gunten et al. (2015) report findings from a retrospective observational study looking at device lifespan. Only 26.3% (n=1,284) of devices included in the study were CRT-Ds, but the results are presented separately for this subgroup. ENDURALIFE-powered CRT-Ds comprised 39% of Boston Scientific devices. Median follow-up was 4.4 years. For devices implanted after 2006, the proportions of devices still working after 6 years were 97.6% (Boston Scientific), 26.5% (St Jude Medical), 46.3% (Medtronic) and 44.9% (Biotronik).
- 3.7 Lau et al. (2015) is a published abstract based on a conference poster presentation reporting the findings from a UK hospital. The study compared battery life after 6 years of use in Boston Scientific ENDURALIFE-powered CRT-Ds, and Medtronic and St Jude Medical CRT-Ds. At 6-year follow-up, none of the Boston Scientific devices needed replacement because of battery depletion. St Jude Medical CRT-Ds first began to reach ERI after 2.8 years, and Medtronic CRT-Ds after 2.5 years. Pairwise comparisons showed a significant difference between Boston Scientific and St Jude Medical (p=0.0018) and between Boston Scientific and Medtronic (p<0.0001).
- 3.8 Williams and Stevenson (2014) is a published abstract from a conference poster presentation reporting battery life of CRT-Ds. The primary end point was device replacement after reaching ERI. A total of 91 CRT-Ds

were implanted from July 2008 to July 2010 (final device follow-up: October 2013): Boston Scientific n=53 (company's submission states that 51 [96.2%] were ENDURALIFE-powered), St Jude Medical n=10 and Medtronic n=28. At 4-year follow-up, the ERI rates were 1.9% (Boston Scientific), 10.0% (St Jude Medical) and 50.0% (Medtronic). Multivariate analysis showed that CRT-Ds reaching ERI had higher right ventricle lead output, left ventricle lead output and right ventricle pulse width (no values reported).

PPRs reporting on device malfunction and survival probability

3.9 Product performance reviews (PPRs) are based on devices that have been replaced and returned to the manufacturer, as well as additional information provided to the manufacturer from various sources about out-of-service devices that have not been returned. They aim to report device malfunctions in a standard format. PPRs report survival probability in 2 ways (based on real, observed data): survival free from both malfunction and normal battery depletion, and survival free of malfunction alone leading to device removal (cases of normal battery depletion are excluded from the analysis). In both cases the definition of 'normal battery depletion' is a function of the manufacturer's predicted device lifespan (based on bench testing, which differs by manufacturer and may not accurately reflect clinical performance). The company presented PPRs from 5 manufacturers of CRT-Ds in its submission. The EAC accepted that the PPRs showed that normal battery depletion, rather than CRT-D malfunction, is the main reason for CRT-D replacement. However, it judged that data in the PPRs could not be used to reliably compare the lifespan of ENDURALIFE-powered devices with that of other devices.

Adverse events associated with CRT-D replacement

3.10 Lewis et al. (2016) is a systematic review assessing the risks and benefits of replacing implantable cardioverter defibrillators, which included 17 studies (n≥167,000 patients). The median rate for major complications was 4.05% (range: 0.55% to 7.37%), of which the most frequent was infection needing antibiotic therapy and/or device removal (median rate 1.70% [range: 0 to 5.23%]). Other reported major

complications included haematoma needing evacuation (median 0.57%; range: 0 to 1.55%), stroke (median 0.45%, range 0.01% to 0.82%) and reoperation for any other reason (such as pocket erosion or device repositioning because of pain; median 1.56%; range: 0.07% to 3.24%). The median rate for minor complications was 3.5% (range: 0.36% to 7.37%), with the most frequent being pocket haematoma (median 0.93%; range: 0.35% to 3.49%). Other reported minor outcomes include: incisional infection (median 0.9%; range: 0.01% to 1.77%) and discomfort or pain at the site (median 0.44%; range: 0.39% to 0.45%).

- 3.11 Polyzos et al. (2015) conducted a systematic review and meta-analysis on risk factors associated with cardiac implantable electronic device infection, including 60 studies with a total of 233,184 patients. The average reported device infection rates were 1.6 for prospective studies (n=21 studies), 1.0% for case-control studies (n=9 studies) and 1.2% for retrospective cohort studies (n=30 studies). The pooled OR for the risk of infection associated with generator change (20 studies; 33,322 patients) was 1.74 (95% CI 1.22 to 2.49). Device replacement or revision was associated with a pooled OR of 1.98 (95% CI 1.46 to 2.70) for infection. The authors concluded that a 'decision to replace a device should be made on a risk/benefit approach weighting the risk for death because of device failure, the rate of device failure, and the risk for procedure-related death'.
- 3.12 Zeitler et al. (2015) present a systematic review and meta-analysis of the complications associated with the replacement of cardiac implantable electronic device generators, following US Food & Drug Administration (FDA) recall. The review included 7 studies (1,435 patients) with a primary end point of major complications and mortality; other end points included reoperation and pocket revision. Device replacement following FDA recall was associated with a combined major complication rate of 2.60% (95% CI 0.9% to 4.8%). Five of the 7 included studies reported mortality, which showed an overall mortality of 0.4% (95% CI 0.1% to 1.1%). The rate of reoperation/pocket revision (5 studies) was 2.7% (95% CI 0.8% to 5.1%). The authors conclude that generator replacement in response to an FDA recall has a similar rate of major complications as elective generator replacement. The authors also conclude that patient and device characteristics, patient preference and remaining battery life

should all be considered when replacing devices, elective or otherwise.

- 3.13 Nichols et al. (2016) investigated the incidence of lead damage following cardiac implantable electronic device replacement procedures and its economic impact. The study included 45,252 patients: 22,557 (50%) pacemaker generator replacements, 20,632 (46%) implantable cardioverter defibrillator replacements, and 2,063 (5%) CRT-D device replacements. Lead damage was observed in 406 patients (0.90%) at a median of 107 days following device replacement. Lead damage incidence was 1.94% for patients with CRT-Ds. In a Cox proportional hazards model, controlling for patient demographic and clinical characteristics, CRT-D replacement showed >2.5 times (HR 2.58, 95% CI 1.73 to 3.83) the risk of lead damage compared with pacemaker replacement. Of the 406 patients with lead damage, 368 (91%) were inpatients with a median length of stay for lead damage of 3 days; this did not significantly differ based on device type. The mean cost of lead damage management across all device types in the first year was \$25,797. Average lead damage hospitalisation costs were significantly different across device types: \$19,959 for pacemaker replacement; \$24,885 for implantable cardioverter defibrillator replacement; and \$46,229 for CRT-D replacement ($p=0.048$). The authors conclude that the higher rates of lead damage observed in implantable cardioverter defibrillator and CRT-D replacement are likely to be attributable to the greater number of and complexity of leads in these procedures.
- 3.14 Lovelock et al. (2014) investigated the risk of lead alerts after replacing implantable cardioverter defibrillators. This study utilised patients enrolled on the ALTITUDE project, an initiative to prospectively analyse data obtained from implanted Boston Scientific devices through its LATITUDE home monitoring system. A total of 60,219 patients were eligible for inclusion in the study, of which 7,458 patients (12.4%) had implantable cardioverter defibrillator replacement. A time-dependent Cox proportional hazards model (adjusted for age, gender and device type) was used to evaluate potential associations between lead failure and device replacement. Lead performance in the 7,458 patients having device replacement was compared with leads of similar age (68 months) in patients who did not have device replacement. Patients who had device replacement showed a 5-times higher lead alert rate (HR 5.20,

95% CI 3.45 to 7.84) compared with those who did not; this was significantly different even when covariates were adjusted for ($p < 0.001$). Younger age and single-lead implantable cardioverter defibrillators were also associated with an increase in lead alerts: HR 1.02, 95% CI 0.98 to 0.99, $p < 0.001$; HR 2.49, 95% CI 1.96 to 3.17, $p < 0.001$ respectively. However, both age and system type were associated with lead alerts to a lesser degree than device replacement. The authors suggest that surveillance is needed after device replacement in addition to technique development and lead modifications to minimise the risk of lead damage during surgery. In another study, Lovelock et al. (2012) reported that the rate of failure in Medtronic Fidelis leads was 20.8% following device replacement and 2.5% in lead age-matched controls ($p < 0.001$).

- 3.15 Kirkfeldt et al. (2014) was a retrospective multicentre (14 hospitals) cohort study in Denmark which analysed complications occurring within 6 months of cardiac electronic devices implanted between May 2010 and April 2011. The analysis included 5,918 patients: 74% ($n=4,355$) had new implants, 19% ($n=1,136$) had device replacements and 7% ($n=427$) had system upgrades or lead revisions. The complication rate was 5.9% following a device replacement. Infection rates for new implants and generator replacements were 0.6% and 1.5% respectively. When complications were categorised, 3.5% of patients experienced a major complication within 6 months of a device replacement.

EAC's critique of the clinical evidence

- 3.16 The EAC felt that although the studies of battery life were done under similar conditions to normal clinical practice, they were retrospective and it was not possible to determine the rationale for choice of CRT-D. It concluded that the published studies demonstrate that ENDURALIFE-powered CRT-Ds implanted between 2008 and 2010 lasted longer than other CRT-Ds implanted during the same period. However, some of the CRT-Ds in these studies, particularly the comparator devices, are no longer marketed.
- 3.17 The EAC accepted the company's submission of evidence on the rate of complications following CRT-D replacements. The EAC acknowledged that the PPRs submitted by the company demonstrated that most

CRT-Ds are replaced because of normal battery depletion, and not device malfunctions.

Committee considerations

- 3.18 The committee concluded that ENDURALIFE-powered CRT-Ds have a greater battery capacity and longer battery life compared with other CRT-Ds available at the time of the published studies. It noted that, because of the follow-up time needed to study battery life, the retrospective, observational studies presented included CRT-Ds no longer marketed. Clinical experts advised the committee that the company's claims relating to battery life and the ENDURALIFE battery technology have been borne out in their own subsequent clinical experience, as well as in the published literature.
- 3.19 The committee heard that other technologies claim to offer similar advantages to ENDURALIFE-powered CRT-Ds, but these have not been reviewed in this assessment.
- 3.20 The committee was advised by clinical experts that in terms of determining the lifespans of different CRT-Ds, published, peer-reviewed clinical studies are a more reliable source of information than unpublished, extrapolated and predicted lifespan data. The committee considered that the updating and publication of further clinical outcome studies in patients with CRT-Ds from all manufacturers would be helpful. In this regard, the committee was made aware of the existence of a large volume of data possessed by the National Institute for Cardiovascular Outcomes Research (NICOR) relating to CRT-Ds implanted in the NHS since 2008 (when ENDURALIFE-powered CRT-Ds entered the market). These data include lifespan outcomes for almost all CRT-Ds implanted in the UK since 2008. The committee was advised that further analyses of these data may provide valuable insights into how long different CRT-Ds last in real-world clinical practice. The committee encouraged the publication of these lifespan outcomes.
- 3.21 The committee heard from clinical experts that battery depletion depends on a number of factors including the needs of the patient, lead technology, battery design and the algorithms used in the CRT-D.

However, it was advised that a central factor in determining device lifespan is the ampere hours a battery can carry. The experts stated that accepting recent developments in battery technology by all CRT-D manufacturers, ENDURALIFE-powered batteries still have one of the largest ampere hours ratings.

- 3.22 The experts acknowledged that advances in CRT-D technology continue to be made by all manufacturers, particularly in minimising battery drain. However, the experts advised that these developments applied to all manufactured devices including ENDURALIFE-powered CRT-D devices.
- 3.23 The committee was advised that replacement procedures are associated with a risk of serious complications and that complications are more common in replacement than primary implants. Infection can have major consequences in terms of patient morbidity and resource use, including the need for hospital admission that may last days or weeks. The committee heard from a patient expert that replacement procedures have a detrimental impact on quality of life. The clinical experts also advised that patients see replacement procedures as a significant life event.
- 3.24 The committee heard from the clinical experts that predicting a patient's individual life expectancy after device implantation is difficult. Nonetheless, experts advised that given the prognostic benefit of CRT-D implantation in patients with heart failure, the choice of a CRT-D with a greater lifespan is logical.
- 3.25 The committee was advised that CRT-Ds differ in size and shape between manufacturers and that Boston Scientific devices are slightly thinner than others. Experts stated that the shape of the CRT-D is sometimes more important than the size, and that the choice of device needs to be personalised to the patient's individual needs. This usually involves shared decision-making between the patient and the clinician.

4 NHS considerations

System impact

- 4.1 The company claimed that fewer avoidable replacement procedures will lead to a reduction in hospital admissions, bed days and procurement costs. Fewer replacement procedures could result in more efficient use of resources, because it would allow more primary (that is, non-replacement) implants within the same resource constraints. The company also claimed that using ENDURALIFE-powered cardiac resynchronisation therapy-defibrillator (CRT-D) devices could lead to cost savings through a reduction in associated costs such as post-operative complications and infections resulting from replacement procedures.

Committee considerations

- 4.2 The committee was advised by clinical experts that although replacements are done for a variety of reasons, including lead failures, 80% to 90% of CRT-D replacements are because of battery depletion.
- 4.3 The committee heard from experts that the cost of replacing a CRT-D is between £10,000 and £15,000, not including the cost of additional leads.
- 4.4 The committee was advised that remote monitoring affects battery drain across all manufacturers' devices to a variable degree.
- 4.5 The committee heard from experts that, despite increased battery life being an important patient benefit, it is standard practice for a single centre to use CRT-Ds from more than 1 manufacturer. The rationale is to spread the risk of undue pressure on clinical services in the face of possible future device-related technical failure necessitating recall and replacement. In view of this important consideration, professional advice was that it was unwise for a department to rely entirely on the use of a CRT-D from a single manufacturer.

5 Cost considerations

Cost evidence

- 5.1 The company identified 7 studies that incorporated a cost-effectiveness analysis. It did not rely on these economic studies for its model, but the structure of the de novo model is similar to that described in Gadler et al. (2016). The external assessment centre (EAC) judged the company's search strategy and inclusion/exclusion criteria reasonable, but noted that it could be improved with access to more databases and a more thorough strategy. The EAC considered that the population used by the company in its selection of economic evidence – 'patients implanted with cardiac resynchronisation therapy-defibrillators (CRT-Ds)' – differed from the population specified in the scope. The company's population is broader and the EAC acknowledged that this probably reflects the lack of detail in the published evidence on the specific criteria used to define heart failure and CRT-D use from the [NICE's technology appraisal guidance on implantable cardioverter defibrillators and cardiac resynchronisation therapy for arrhythmias and heart failure](#).
- 5.2 The EAC excluded 3 studies included by the company because they were outside the scope. Boriani et al. (2013) report on a model comparing hypothesised CRT-Ds with 4-year and 7-year lifespans over a 15-year time horizon. The devices were not specifically named technologies and the lifespans were not based on data, but were chosen specifically to investigate how battery life affects costs. Biffi et al. (2011) focused on implantable cardioverter defibrillators and included only 10 patients with CRT-Ds. It did not include devices from Boston Scientific. The Chung et al. (2015) abstract does not directly compare specific devices although it includes a device survival curve based on manufacturer data, but looks at the costs for different patient groups using devices with different lifespans.
- 5.3 Gadler et al. (2016) describes an economic model with a 6-year time horizon based on the data from Landolina (2015; see [section 3.5](#)), Swedish ICD and Pacemaker Registry and Swedish public tendering data.

Survival data were taken from Yao et al. (2007). The authors found in the base case that using Boston Scientific devices reduced replacement procedures and saved SEK 19,172 (£1,687 based on XE.com currency conversion on 15 July 2016) per patient over 6 years. The study was funded by Boston Scientific.

- 5.4 Landolina et al. (2016) is an economic analysis based on a subset of the data from Landolina et al. (2015), with a 6-year time horizon and 2 perspectives: a hospital perspective and the Italian healthcare system perspective. Boston Scientific provided funding for the economic analysis. Of 1,726 heart failure patients in Landolina 2015, 1,399 were included in the economic analysis. The analysis compared devices released between 2007 and 2010 with devices released between 2003 and 2007 for 3 manufacturers (Boston Scientific, Medtronic and St Jude Medical) and for all manufacturers together. Weighted average prices of the devices were taken from tender information. The authors found that among recent-generation CRT-Ds from different manufacturers, the total cost per patient over 6 years ranged from €25,579 to €31,536 (£21,665 to £26,711 based on XE.com currency conversion on 12 July 2016), with a maximum difference in cost of 40% for hospitals and 19% for the Italian healthcare system.
- 5.5 Priest et al. (2015) is a published abstract from a conference poster presentation comparing the costs for industry-standard and longer-lifespan devices from an Australian health system perspective over 15 years, using real-world data for implantable cardioverter defibrillators and CRT-Ds (using the methods described by Boriani et al. 2013). Patient survival following first implant was taken from published literature. Average battery life was taken from a recent NICE review (not specified, but the figures quoted are found in [NICE's technology appraisal guidance on implantable cardioverter defibrillators and cardiac resynchronisation therapy for arrhythmias and heart failure](#)), and Boston Scientific real-world battery life data from more than 100,000 for implantable cardioverter defibrillators using the LATITUDE NXT remote monitoring system. The study concluded that if all patients switched from industry-standard devices to longer-lifespan batteries, the average cost per patient would fall by 19% and overall number of replacements would fall by 70%. This would result in cumulative cost savings of more than

\$900 million over 15 years.

- 5.6 The paper by Duxbury et al. (2014) is a published abstract from a conference poster presentation reporting the economic impact of implanting cardiac devices with longer lifespans from a UK perspective. The methodology was similar to that of the Priest et al. study (2015), in that it was based on Boriani et al. (2013). It also used the average lifespans described in the NICE technology appraisal and Boston Scientific real-world battery life data using the LATITUDE NXT remote monitoring system. The authors modelled the potential cumulative costs over 10 years for industry-standard and longer-lifespan devices using real-world battery data for implantable cardioverter defibrillators and CRT-Ds. The study concluded that using devices with longer battery life could result in cumulative savings of up to £158 million over 10 years.

EAC's critique of the cost evidence

- 5.7 The EAC identified that the main weakness of the published economic evidence was it relates to devices no longer marketed, because of the rapid turnover of new models of the technology. The study by Gadler et al. (2016) was funded by Boston Scientific, so may be subject to bias. The EAC considered the lifespan data for the LATITUDE NXT system used in Priest et al. (2015) may not be directly comparable with that reported in NICE's technology appraisal guidance on implantable cardioverter defibrillators and cardiac resynchronisation therapy for arrhythmias and heart failure, because the patient populations may be different. The EAC concluded that because the Priest et al. (2015) and Duxbury et al. (2014) studies are only available as abstracts, the results should be treated with caution.

Cost model

- 5.8 The company presented a de novo economic model adapted from Gadler et al. (2016) estimating mean cost savings per patient. The model is a decision tree with a 6-year time horizon and an NHS perspective. It compares Boston Scientific ENDURALIFE-powered CRT-Ds with Medtronic and St Jude Medical CRT-Ds. For each device there are branches for procedural complications or no complications, with further

branches for death, replacement or no replacement at 1 year and at each subsequent year. Clinical data in the model are taken from the Landolina et al. (2016) study on event-free battery survival and Yao et al. (2007) for cumulative probability of patient survival. The incidence of complications is taken from Tang et al. (2010) and the follow-up arrangements from NHS England 2013/14 NHS standard contract for cardiology: implantable cardioverter defibrillator and cardiac resynchronisation therapy (adult). The model assumes follow-up appointments at 6-month intervals with an additional post-procedure appointment.

- 5.9 The company's scenario analyses included exploring differences in device survival and device cost to identify thresholds at which the model becomes cost neutral. The price was varied by $\pm 20\%$ for each device separately using a one-way sensitivity analysis. The analyses showed that the cost model is highly sensitive to changes in both device survival and device cost. Higher device survival resulted in a marked decrease in relative costs. The one-way sensitivity analysis of device cost showed that ENDURALIFE-powered CRT-Ds remained cost saving.
- 5.10 The company's base case showed that ENDURALIFE-powered CRT-Ds cost £22,322 per patient over a 6-year period compared with £27,309 and £29,158 per patient for St Jude Medical and Medtronic CRT-Ds respectively. The company therefore estimated that using ENDURALIFE-powered CRT-Ds would save between £4,987 and £6,836 per patient over 6 years. Cost savings come mainly from fewer replacement procedures.

Additional work by the external assessment centre

- 5.11 The EAC re-ran the company's base case and univariate sensitivity analyses and conducted additional analyses using its preferred estimates. The EAC also did a threshold analysis using the average selling price for ENDURALIFE-powered CRT-Ds and allowing the cost of the comparator devices to fall to the point at which each becomes cost neutral. The main changes to the company's model were:
- Changes to the list prices of ENDURALIFE-powered CRT-Ds and both comparators.

- Using warranty data from the comparator manufacturers instead of that from Boston Scientific.
 - Using NHS reference costs instead of Payment-by-Results tariff costs.
 - Changes to the sensitivity analysis for complication rates (infection), based on the results of a large Danish cohort study (Kirkfeldt et al. 2014). This changed the infection rate from 2.4% to 0.6% for new implants.
 - Using patient survival data from the National Institute for Cardiovascular Outcomes Research (NICOR) instead of from Yao et al. (2007).
- 5.12 The results of the EAC analysis suggested that changing the device cost in the model to the lowest and highest list price for each of the 3 manufacturers results in ENDURALIFE-powered CRT-Ds becoming more costly than those from Medtronic, but remaining cost saving compared with those from St Jude Medical.
- 5.13 The threshold analysis investigated the effect of introducing a price difference between the devices, and calculated the threshold at which ENDURALIFE-powered CRT-Ds become cost incurring compared with the comparators. The results showed that, using the same cost of implanting and replacing the CRT-D as used in the company's base case, ENDURALIFE-powered CRT-Ds become cost incurring when they are £4,858 more expensive to purchase than Medtronic CRT-Ds and £3,858 more expensive to purchase than St Jude Medical CRT-Ds, with all other model inputs unchanged.
- 5.14 Using NHS reference costs instead of the Payment-by-Results tariff increased the cost of ENDURALIFE-powered CRT-Ds from £22,322 in the company's base case to £30,957. ENDURALIFE-powered CRT-Ds remain cost saving compared with the comparators but to a lesser extent than in the company's base case.
- 5.15 Substituting the actual warranty information supplied by the manufacturers into the model showed that ENDURALIFE-powered CRT-Ds remained cost saving.
- 5.16 Changing the rate of infection for new implants from 2.4% to 0.6% had little effect on the costs.

- 5.17 Following expert advice, the EAC contacted NICOR, which holds a registry of NHS patients who have had CRT-Ds implanted, including data on overall survival.
- 5.18 Using patient survival data from NICOR to replace that from Yao et al. (2007), ENDURALIFE-powered CRT-Ds remained cost saving when using the company's base-case device cost. At the lowest and highest list prices, ENDURALIFE-powered CRT-Ds become more costly than those from Medtronic, but remain cost saving compared with those from St Jude Medical.
- 5.19 The EAC concluded that the main driver of the cost model was device price.
- 5.20 The committee considered that the 6-year time horizon made the cost case uncertain. The EAC was therefore asked to carry out further analyses extrapolating the data available over a patient's lifetime (sections 5.20 to 5.24).
- 5.21 NICOR provided unpublished data in confidence which showed patient survival by age group after primary implantation of a CRT-D. The EAC extrapolated CRT-D lifespan to 15 years using a survival profile for comparator devices: this took an average distribution based on Medtronic and St Jude Medical CRT-D lifespans reported in Landolina et al. (2015), and then applied the average distribution to the ENDURALIFE-powered CRT-Ds from the point at which the ENDURALIFE-powered CRT-Ds begin to reach the elective replacement indicator (ERI), at 5 years following implantation.
- 5.22 The EAC extrapolated patient survival to 15 years using NICOR data for patients aged 50 to 84 years at primary implantation.
- 5.23 Using the average selling price in the company's base case and the extrapolated data outlined above, the results showed that ENDURALIFE-powered CRT-Ds cost £28,234 per patient over 15 years compared with £30,354 and £33,861 per patient for St Jude Medical and Medtronic CRT-Ds respectively. The EAC therefore estimated that using ENDURALIFE-powered CRT-Ds could save between £2,120 and £5,627

per patient over 15 years.

- 5.24 A threshold analysis investigated the effect of allowing a price difference between the devices, and calculated the threshold at which ENDURALIFE-powered CRT-Ds become cost incurring compared with the comparators. The results showed that, using the same cost of implanting and replacing the CRT-D as used in the company's base case, ENDURALIFE-powered CRT-Ds become cost incurring when they are £3,304 more expensive to purchase than Medtronic CRT-Ds and £1,404 more expensive to purchase than St Jude Medical CRT-Ds.

EAC's critique of the cost model

- 5.25 The EAC considered that the 6-year time horizon used in the model may overestimate the potential cost saving of a slightly longer-lasting device. It concluded that a time horizon over the patient's lifetime may be more appropriate.

Committee considerations

- 5.26 The committee was advised that device costs were accurately reflected in the company's base case, which used average selling prices, and that prices are similar between manufacturers. List prices are not a true reflection of what the NHS pays for CRT-Ds.
- 5.27 The committee concluded that it would be difficult to ascertain actual NHS device costs for ENDURALIFE-powered and comparator CRT-Ds. The EAC was asked to carry out a differential cost threshold analysis to overcome some of these uncertainties.
- 5.28 The committee accepted that the EAC's revisions to the company's cost modelling provided the most plausible estimates for the cost consequences of adopting ENDURALIFE-powered CRT-Ds.
- 5.29 For the guidance review, the external assessment centre revised the model to reflect 2020 costs (original guidance costs are given in brackets). The main parameter changes were the estimated costs:

- ENDURALIFE-powered CRT-Ds £15,322 (£12,404)
- procedural complications
 - £24,686 (£21,774) for infection
 - £6,975 (£6,152) for a complication requiring reintervention
 - £20,418 (£18,010) for device pocket issues
- de novo procedures £5,931 (NHS tariff costs; EY01B) and £3,342 (NHS reference costs; EY01B) (£6,201; EY10B)
- replacement procedures £5,931 (NHS tariff costs; EY01B) and £3,342 (NHS reference costs; EY01B) (£4,700)
- follow-up procedures £78 (NHS tariff costs) and £135 (NHS reference costs) (£96; WF01A).

Base-case results for the 2020 revised model shows the cost saving associated with ENDURALIFE-powered CRT-Ds was between £2,614 and £6,941 per patient using NHS national tariff costs and between £2,313 and £6,140 per patient using NHS reference costs over 15 years. Further details of the 2020 revised model are in the [revised model summary in the review decision \[2021\]](#).

6 Conclusions

- 6.1 The committee concluded that there is good evidence to support the clinical benefit of longer battery life and the associated reduction in cardiac resynchronisation therapy-defibrillator (CRT-D) replacements.
- 6.2 The committee concluded that developments in CRT-D technology are ongoing and that the evidence available suggests that the advantages of longer battery life have not been surpassed by other types of technical advances.
- 6.3 The committee noted the potential hazards of a clinical service relying solely on a single manufacturer's CRT-D devices.
- 6.4 The committee encouraged further studies that provide data on the battery life of different CRT-Ds, including an analysis of currently available UK NHS clinical data.
- 6.5 Based on cost modelling, the committee concluded that using ENDURALIFE-powered CRT-Ds in patients with heart failure is likely to save costs by reducing the number of replacement procedures.

7 Committee members and NICE project team

Committee members

This topic was considered by the medical technology advisory committee which is a standing advisory committee of NICE.

Committee members are asked to declare any interests in the technology to be appraised. If it is considered there is a conflict of interest, the member is excluded from participating further in that evaluation.

The minutes of each committee meeting, which include the names of the members who attended and their declarations of interests, are posted on the NICE website.

NICE project team

Each medical technology appraisal is assigned to a team consisting of 1 or more health technology analysts (who act as technical leads for the appraisal) and a technical adviser.

Liesl Millar

Technical analyst

Paul Dimmock

Technical analyst (evaluations)

Update information

July 2021: We updated this guidance to reflect 2020 costs and revise cost-saving estimates. These are marked [2021]. Details of the changes are explained in the [review decision](#).

ISBN: 978-1-4731-2379-3

Accreditation

